

IEC SYSTEM FOR CONFORMITY TESTING AND  
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION  
DES EQUIPEMENTS ELECTRIQUES (IECEE)  
METHODE OC

## CB TEST CERTIFICATE

## CERTIFICAT D'ESSAI OC

Product  
Produit

Switching Power Supply, Built-In AC/DC

Name and address of the applicant  
Nom et adresse du demandeur

GLOBTEK INC  
186 VETERANS DR  
NORTHVALE NJ 07647, USA

Name and address of the manufacturer  
Nom et adresse du fabricant

GLOBTEK INC  
186 VETERANS DR  
NORTHVALE NJ 07647, USA

Name and address of the factory  
Nom et adresse de l'usine

See Information on page 2

Rating and principal characteristics  
Valeurs nominales et caractéristiques principales

See Information on page 2

Trademark (if any)  
Marque de fabrique (si elle existe)



Globtek or  
N/A

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

GTM9200P Series

Model / Type Ref.  
Ref. de type

Additional information (if necessary)  
Information complémentaire (si nécessaire)

This report comprises 2 enclosures.

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

**IEC 60601-1 (1988) Second Edition,  
with Amendment No. 1 (1991) and No. 2 (1995)** with the exception of:  
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and  
Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common  
Modifications. Additionally evaluated to EN 60601-1: 1990 + A1:1993 + A2:1995.  
See Test Report for National Differences.  
E172861-A15-CB-1

as shown in the Test Report Ref. No.  
which forms part of this Certificate  
comme indiqué dans le Rapport d'essais numéro  
de référence qui constitue partie de ce Certificat

**This CB Test Certificate is issued by the National Certification Body**  
**Ce Certificat d'essai OC est établi par l'Organisme National de Certification**




**Underwriters  
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008, FAX INT\* +1 847 313 3008  
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2010 November 29

Signature:



Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND  
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION  
DES EQUIPEMENTS ELECTRIQUES (IECEE)  
METHODE OC

## CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Factories:

GLOBTEK INC  
186 VETERANS DR  
NORTHVALE NJ 07647, USA  
GLOBTEK (SUZHOU) CO LTD  
BLDG 4, #76  
JINLING EAST RD, SUZHOU PARK  
SUZHOU, JIANGSU 215021, CHINA

Ratings:

Input -

Voltage: 100-240 Vac

Frequency: 50-60 Hz

Rated Current for 350 W units: 5.0 A  
(except for GTM9200P2313.3 which is rated at 4.0 A)

Rated Current for 200 W units: 3.2 A  
(except for Models GTM9200P1323.3 which is rated 2.5 A)

Note: See the Model Differences section of report for Output ratings.

Additional information (if necessary)  
*Information complémentaire (si nécessaire)*

This CB Test Certificate is issued by the National Certification Body  
*Ce Certificat d'essai OC est établi par l'Organisme National de Certification*



**Underwriters  
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US  
333 Pfingsten Road, Northbrook, IL 60062-2096  
United States of America  
TEL INT\* +1 847 664 3008 FAX INT\* +1 847 313 3008  
email: [jolanta.m.wroblewska@us.ul.com](mailto:jolanta.m.wroblewska@us.ul.com)

Date: Issued: 2010 November 29

Signature:



Jolanta M. Wroblewska