

## **COVER PAGE FOR TEST REPORT**

Product Category:	Power Supplies, Medical and Dental
Product Category CCN:	QQHM2, QQHM8
Complementary Product Categories:	Power Supplies for Information Technology Equipment Including Electrical Business Equipment(QQGQ2, QQGQ8)
Test Procedure:	Component Recognition
Product:	Switching Power Supply
Model/Type Reference:	GTMB1057-6010, -6012, -6015, -6018, -6019, -6020, -6022, -6024, -6030.  Model numbers may be followed by an additional suffix X.X. Where X.X. represents the output deviation from the standard model; subtract X.X. from the standard output voltage to obtain the intended output voltage.
Rating(s):	Input: 100-240 VAC, 50-60Hz, 1.5A  Output: 10-30 VDC, 2-6 A, 60W
Standards:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 1st Edition, 2003-11 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
Applicant Name and Address:	GLOBTEK INC 186 VETERANS DR NORTHVALE NJ 07647
This Report includes the following parts, in addition to this cover page:	
<ol style="list-style-type: none"><li>1. Specific Technical Criteria</li><li>2. Clause Verdicts</li><li>3. Enclosures</li></ol>	


This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Test Report By:



Maximiliano S. Rodriguez  
Project Engineer  
Underwriters Laboratories Inc.

Reviewed By:



David V. Alma  
Staff Engineer  
Underwriters Laboratories Inc.