




Reference No. : WTX23X10233369E001
 Applicant : GlobTek, Inc.
 Address : 186 Veterans Dr. Northvale, NJ 07647 USA
 Manufacturer : 1: GlobTek, Inc. 2: GlobTek (Suzhou) Co., Ltd
 1: 186 Veterans Dr. Northvale, NJ 07647 USA
 Address : 2: Building 4, No. 76, Jin Ling East Rd., Suzhou
 Industrial Park, Suzhou, JiangSu 215021, China
 Product Name : ITE/Medical Power Supply
 Trade Name : 
 Model(s) : See Model List
 Ratings : Refer to the test report

Test Standards:

EN 55032:2015+A1:2020; EN 55035:2017+A11:2020;
 EN IEC 61000-3-2:2019+A1:2021; EN 61000-3-3:2013+A2:2021

The above product has been tested by us with the listed standards and found in compliance with the European Electromagnetic Compatibility Directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

EN 55032: Electromagnetic compatibility of multimedia equipment - Emission requirements
EN 55035: Electromagnetic compatibility of multimedia equipment - Immunity requirements
EN IEC 61000-3-2: Electromagnetic Compatibility (EMC) Part 3-2: Limits for harmonic current emissions (Equipment input current up to and including 16A per phase)
EN 61000-3-3: Electromagnetic Compatibility (EMC) Part 3-3: Limitation of voltage supply systems for equipment with rated current $\leq 16A$

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the conformity declaration, the CE marking as shown below can be affixed on the equipment.



The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308
 Http://www.waltek.com.cn

Fax: +86-755-3366 3309
 E-mail: sem@waltek.com.cn



ANNEX 1 TO ADDING MODEL

Certificate No. : WTX23X10233369E001

Model List

GT(M or -)21097-X-Y.Y series

GT(M or -)21097-X-Y.Y

Where "M" or "-" for market identification and not related to safety.

"X" denotes the standard output power and voltage, which can be 2003, 3005, 4509, 5012, 5015, 5018, 5024 or 5048, the former two numbers of figures indicate max. output power, 50W max. and the latter two numbers of figures indicate standard model output voltage.

"-Y.Y" is optional variable or blank for specifying output voltage deviation from standard model; which means subcontracting volts from standard output voltage in 0.1V increments or blank; 0.1min.to 23.9 max., actual voltage range is 3.3-48V only.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn