


Reference No. : WTX21X09095476E
Holder : GlobTek, Inc.
Address : 186 Veterans Dr. Northvale, NJ 07647 USA
Product : Medical/ITE Power Supply
Trade Name : 
Model(s) : See Model List
Ratings : Refer to the test report



Test Standards:

EN 55032:2015+A11:2020
EN 55035:2017+A11:2020
EN IEC 61000-3-2:2019
EN 61000-3-3:2013+A1:2019

The above product has been tested by us with the listed standards and found in compliance with the European Electromagnetic Compatibility Directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

EN 55032: Electromagnetic compatibility of multimedia equipment - Emission requirements
EN 55035: Electromagnetic compatibility of multimedia equipment - Immunity requirements
EN IEC 61000-3-2: Electromagnetic Compatibility (EMC) Part 3-2: Limits for harmonic current emissions (Equipment input current up to and including 16A per phase)
EN 61000-3-3: Electromagnetic Compatibility (EMC) Part 3-3: Limitation of voltage supply systems for equipment with rated current $\leq 16A$

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the conformity declaration, the CE marking as shown below can be affixed on the equipment.



The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308
Http://www.waltek.com.cn

Fax: +86-755-3366 3309
E-mail: sem@waltek.com.cn

ANNEX 1 TO ADDING MODEL



Certificate No. : WTX21X09095476E

Model List

GT-46600-6012-T2, GT-46600-6524-T3

GT-46600-WWVV-X.X-TB*

WW is the standard output wattage, with a maximum value of "65",

VV is the standard rated output voltage designation, with a value of "12" "15" and "24";

-X.X denote the output voltage differentiator, subtracting X.X volts from standard output voltage VV in 0.1V increments, the actual output voltage range is 12-24V, blank is to indicate the no voltage different.

B can be 2 or 3 or 3A, 2 means C8 inlet type, 3 means C14 inlet type, 3A means C6 inlet type

The last * denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Http://www.waltek.com.cn

Fax: +86-755-3366 3309

E-mail: sem@waltek.com.cn