



# TEST CERTIFICATE

This certificate does not include the authorization to use the GS-mark  
(= Tested Safety).

Application No.: 2017	Application Date: 2000-11-28	Expert: Mr. Neumaier	
Description of test object:	Power supplies for medical equipment		
Type:	GT500148-12, GT500148-15, GT500148-24, Trade name: Glob Tek, Inc. Input voltage 100-240 V~, 50/60Hz, 1.0A max; Class I See page 2 for output ratings.		
Manufacturer's No.:	Prototypes		
Applicant: (Bearer of certificate)	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647 USA		
Manufacturer	s.a.		
Production plant:	Glob Tek, Inc. 186 Veterans Dr. Northvale, NJ 07647 USA	Yang Ban Jing Industrial Zone Lou Feng Exempler Area Suzhou Industrial Park Suzhou Jiangsu CHINA	Pros. Zone P.O. Box 266-110, Kaoshiung 806, TAIWAN R.O.C.

A design model of the above-named equipment of the type as described in the Nemko test report no2000 43 151 dated 2000-11-02 was tested.

Measurements and tests were performed on a test sample in operation, according to the following standards:

**EN 60601-1:90 + A1:93 + A2:95 + A13:96**

The electrical appliance complies with the standard requirements.

The applicant is allowed to affix the safety label "Bauart geprüft" of the Test Institut to the equipment which is identical with the tested model. This certificate is valid till November 29, 2003.

EL/2017/0011



ERG - Elektrotechnische  
Revisionsgesellschaft mbH & Co.  
Test and Certification Institute

  
Dr.-Ing. M. Petirsch  
Head of Certification Institute

Pfanztal, 2000-11-30

Output rating:

Model	Rating
GT500148-24	24 VDC output, 2,1 A maximum
GT500148-15	15 VDC output, 3,3 A maximum
GT500148-12	12 VDC output, 4,0 A maximum

EL/2017/0011



ERG - Elektrotechnische  
Revisionsgesellschaft mbH & Co.  
Test and Certification Institute

Dr.-Ing. M. Petirsch  
Head of Certification Institute

Pfingsttal, 2000-11-30