



Ref. Certif. No.

**SE-90070**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR  
ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product

Multi-Unit Charger

Name and address of the applicant

GlobTek, Inc.  
186 Veterans Drive, Northvale, NJ 07647, USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

*Note: When more than one factory, please report on page 2*

See page 2

Ratings and principal characteristics

See page 2

Trademark (if any)

**Dräger**

Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT-93600SHG3050

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60335-1:2010+A1+A2  
IEC 60335-2-29:2016

As shown in the Test Report Ref. No. which forms part of this Certificate

171201041SHA-001

This CB Test Certificate is issued by the National Certification Body

**Intertek Semko AB**  
Box 1103  
SE-164 22 Kista, Sweden  
Int +46 8 750 00 00

Date: 5 March 2018

**Intertek**

Signature:

Gary Hu





Ref. Certif. No.

**SE-90070**

**Factories**

GlobTek, Inc.  
186 Veterans Drive, Northvale, NJ 07647, USA

GlobTek (Suzhou) Co., Ltd.  
Building 4, No.76 JinLing East Road, Suzhou Industrial Park, Suzhou JiangSu, 215021, China

**Ratings and principal characteristics**

Class I, IP30  
Input: 100-240V~, 50-60Hz, 1.5A,  
Output: 9-12.6VDC, 4A  
AC inlet: Max. 10A,  
AC outlet: Max. 8.5A

**Additional information (if necessary)**

Group differences for CENELEC countries are considered.

Date: 5 March 2018

Signature: 