

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

DC/DC Converter

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
United States of America

Name and address of the manufacturer

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
United States of America

Name and address of the factory

*Note: When more than one factory, please report on
page 2ff.*GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
United States of America☐ See page 2ff.

Ratings and principal characteristics

☐ See page 2ff.

Trademark / Brand (if any)



(GlobTek, Inc.)

Customer Test Facility (CTF) Stage used

Model / Type Ref.

☐ See page 2ff.Additional information (if necessary may
also be reported on page 2ff.)

For national differences refer to test report/s.

A sample of the product was tested and
found to be in conformity withIEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012
IEC 60601-1:2005/AMD2:2020
IEC 60601-1-6:2010
IEC 60601-1-6:2010/AMD1:2013
IEC 60601-1-6:2010/AMD2:2020
IEC 60601-1-12:2014
IEC 60601-1-12:2014/AMD1:2020
EFSH26030113-IE-01-L01,
EFSH26030113-IE-01-L02,
EFSH26030113-IE-01-L04As shown in the Test Report Ref. No.
which forms part of this Certificate

This CB Test Certificate is issued by the National Certification Body



eurofins

Eurofins Product Service GmbH
Storkower Strasse 38C
DE-15526 Reichenwalde b. Berlin
Germany

Date: 2026-09-18

Signature: Max L...



Additional information

Name and address of the factory

GlobTek (Suzhou) Co., Ltd.
Building 4, No. 76 JinLing East Road, Suzhou Industrial Park, Suzhou, JiangSu,
215021
China
AMETEK Reynosa TSE, Inc
Avenida Los Encinos Manzana #1, Lot 3, Parque Industrial Villa Florida, Reynosa,
Tamaulipas C.P. 88730
Mexico

Ratings and principal characteristics

For GTD9X1002-****, GTD9X1003-****
Input: 12-28Vdc, 12.0A
Output: 12.0-54.0Vdc, Max. 8.33A, Max. 100W
For 8016-000102
Input: 12-28Vdc, 12.0A
Output: 15.0Vdc, 4.0A, 120W(Peak)

Model / Type Ref.

GTD9X1002-****, GTD9X1003-****, 8016-000102
The 1st * can be "001" to "100", denotes the rated output wattage designation from 1W to 100W, with a step of 1, which denotes 1W.
The 2nd * can be "12" to "54" or "12.0" to "54.0", denotes the standard rated output voltage designation from 12.0V to 54.0Vdc, with a step of 0.1, which denotes 0.1V.
The 3th * denotes the input structure,
 =-TW denotes input with wires,
 =-TWC denotes input with wires and a cigarette lighter plug installed
The last * denotes any six characters, which can be 0-9 or A-Z or () or – or blank for marketing purposes.

This CB Test Certificate is issued by the National Certification Body



Eurofins Product Service GmbH
Storkower Strasse 38C
DE-15526 Reichenwalde b. Berlin
Germany

Date: 2026-09-18

Signature: Max L...

