Ref. Certif. No.

SI-11646

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Desk-top Power supply for Medical use
Name and address of the applicant	GlobTek, Inc. 186 Veterans Drive Northvale, NJ 07647, USA
Name and address of the manufacturer	GlobTek, Inc. 186 Veterans Drive Northvale, NJ 07647, USA
Name and address of the factory	GlobTek, Inc. 186 Veterans Drive Northvale, NJ 07647, USA
Note: When more than one factory, please report on page 2	☒ Additional Information on page 2
Ratings and principal characteristics	Input: 100-240 V a.c.; 50-60 Hz; 1,6 Amax.; Class I Output: See page 2.
Trademark / Brand (if any)	GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	/
Model / Type Ref.	GTM21097-X-Y.Y series X=2003, 3005, 4509, 5012, 5015, 5018, 5024 or 5048 Y.Y=0.1V min. to 23.9V max or blank See page 2.
Additional information (if necessary may also be reported on page 2)	See the test report for information about national differences. ☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 + A1:2012 + A2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	T223-0536/24 (2024-09-30)

This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Mašera-Spasićeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services (SIST EN ISO/IEC 17065).

Date: 2024-09-30

Signature: Matej Žontar

Name and address of the factory:

1) GlobTek, Inc.
186 Veterans Drive Northvale, NJ 07647, USA

2) GlobTek (Suzhou) Co., Ltd.
Building 4, No. 76, Jinling East Road, Suzhou Industrial Park,
JiangSu 215021, China

Ratings and principal characteristics:

Input: 100-240 V a.c.; 50-60 Hz; 1,6 Amax.; Class I
Output: see table below:

Model	Standard model	Standard model output voltage (Vdc)	Standard model rated output current (A)	Y.Y scope	Output voltage range(V)	Max output power (W)
GTM21097-2003-Y.Y	GTM21097-2003	3.3	6	0.1-0.3 or blank	3-3,3	20
GTM21097-3005-Y.Y	GTM21097-3005	5	6	0.1-1.6 or blank	3,4-5	30
GTM21097-4509-Y.Y	GTM21097-4509	9	5	0.1-3.9 or blank	5,1-9	45
GTM21097-5012-Y.Y	GTM21097-5012	12	4,17	0.1-2.9 or blank	9,1-12	50
GTM21097-5015-Y.Y	GTM21097-5015	15	3,3	0.1-6.9 or blank	12,1-15	50
GTM21097-5018-Y.Y	GTM21097-5018	18	2,8	0.1-6.9 or blank	15,1-18	50
GTM21097-5024-Y.Y	GTM21097-5024	24	2,1	0.1-5.9 or blank	18,1-24	50
GTM21097-5048-Y.Y	GTM21097-5048	48	1,0	0.1-23.9 or blank	24,1-48	50

Additional information (if necessary)

This CB Test Certificate substitutes previously issued CB Test Certificate No. SI-7472, issued on 2019-11-04, due to modification of the test report and standard update.

Date: 2024-09-30

Signature: Matej Žontar