



Shenzhen SEM.Test Technology Co., Ltd.

1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,
Bao'an District, Shenzhen, P.R.C. (518101)

CERTIFICATE OF CONFORMITY

Certificate No.: WTX19X10069291E

The following product has been tested by Shenzhen SEM Test Technology Co., Ltd. with the listing standards and found in conformity with the **EU Directive 2014/30/EU**. It is possible to use CE marking to demonstrate the conformity with this **EMC Directive**.

Report No. : WTX19X10069291E

Applicant : GlobTek, Inc.
Address : 186 Veterans Dr. Northvale, NJ 07647 USA
Manufacturer : GlobTek, Inc.
2. GlobTek (Suzhou) Co., Ltd
Address : 1.186 Veterans Dr. Northvale, NJ 07647 USA
2. Building 4, No. 76, Jin Ling East Rd., Suzhou
Industrial Park, Suzhou, JiangSu 215021, China

Description of Product : ITE/Medical Power Supply
Model No. : GT(M or -) 21097-X-Y.Y series (Type of attachment)


Trade Name : 

Test Standards : **EN 55032:2015/AC:2016-07**
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 55035:2017
EN 60601-1-2:2015

The referred test report(s) show that the product complies with the essential requirements in the above listed standards. The applicant is authorized to use this certificate in connection with the EU declaration of conformity according to Annex 1 of the Directive.



Test Laboratory



JandySo
Supervisor

Date of Issue: Oct. 15, 2019



This certificate of conformity is based on a single evaluation of the submitted sample(s) of the abovementioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.: +86-755-33663309

E-mail: sem@semtest.com.cn Website: www.semtest.com.cn



Shenzhen SEM.Test Technology Co., Ltd.

1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,

Bao'an District, Shenzhen, P.R.C. (518101)

ANNEX 1 TO ADDING MODEL

Certificate No. : WTX19X10069291E

Model List

GT(M or -) 21097-X-Y.Y

Where “M” or “-” for market identification and not related to safety.

“X” denotes the standard output power and voltage, which can be 2003, 3005, 4509, 5012, 5015, 5018, 5024 or 5048, the former two numbers of figures indicate max. output power, 50W max. and the latter two numbers of figures indicate standard model output voltage.

“-Y.Y” is optional variable or blank for specifying output voltage deviation from standard model; which means subcontracting volts from standard output voltage in 0.1V increments or blank; 0.1min.to 23.9 max., actual voltage range is 3.3-48V only.

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the abovementioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.:+86-755-33663309

E-mail: sem@semtest.com.cn Website: www.semtest.com.cn