


Reference No. : WTX23X10233369E002  
Applicant : GlobTek, Inc.  
Address : 186 Veterans Dr. Northvale, NJ 07647 USA  
Manufacturer : 1: GlobTek, Inc.2: GlobTek (Suzhou) Co., Ltd  
1: 186 Veterans Dr. Northvale, NJ 07647 USA  
Address : 2: Building 4, No. 76, Jin Ling East Rd., Suzhou  
Industrial Park, Suzhou, JiangSu 215021, China  
Product Name : ITE/Medical Power Supply  
Trade Name :   
Model(s) : See Model List  
Ratings : Refer to the test report



### Test Standards:

EN 60601-1-2:2015+A1:2021

EN 60601-1-2: Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests



Jandy So / GM  
Date: 2023-12-11

The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn



## ANNEX 1 TO ADDING MODEL

Certificate No. : WTX23X10233369E002

### Model List

GT(M or -)21097-X-Y.Y series

GT(M or -)21097-X-Y.Y

Where "M" or "-" for market identification and not related to safety.

"X" denotes the standard output power and voltage, which can be 2003, 3005, 4509, 5012, 5015, 5018, 5024 or 5048, the former two numbers of figures indicate max. output power, 50W max. and the latter two numbers of figures indicate standard model output voltage.

"-Y.Y" is optional variable or blank for specifying output voltage deviation from standard model; which means subcontracting volts from standard output voltage in 0.1V increments or blank; 0.1min.to 23.9 max., actual voltage range is 3.3-48V only.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn