SLOVENSKA AKREDITACIJA
SIST EN ISO/IEC 17065
CP-001

Licence
for the use of SIQ certification mark

Number:	**SI-SIQ 063/098**	Project file: C20241966
Product:	**Desk-top Power supply for Medical use**	
Type reference:	GTM21097-X-Y.Y series X=2003, 3005, 4509, 5012, 5015, 5018, 5024 or 5048 Y.Y=0.1V min. to 23.9V max or blank	
Trademark:	GlobTek, Inc.	
Ratings:	Input: 100-240 V a.c.; 50-60 Hz; 1,6 Amax.; Class I Output: See page 2.	
Applicant:	**GlobTek, Inc.** **186 Veterans Drive Northvale, NJ 07647, USA**	
Manufacturer:	**GlobTek, Inc.** **186 Veterans Drive Northvale, NJ 07647, USA**	

SIQ hereby grants the right to use the SIQ certification mark of conformity on the products specified in this document. The SIQ certification mark of conformity signifies the compliance of the products with requirements of cited standards.

Certification mark:

Standard:	EN 60601-1:2006 + A1:2013 + A12:2014 + A2:2021
Test report:	T223-0536/24 (2024-09-30)
Remarks:	This licence has been issued under the presumption and conditional on the fact that the licensee holds all necessary legal rights with regard to the product presented for testing and certification. This licence is valid as long as the conditions laid down in the listed standards are not modified significantly and until the licensee complies with the SIQ's rules on product certification.

Date: 2024-09-30

Authorized signature: Matej Žontar

Only integral publication of this licence is allowed. This licence may only be reproduced in its entirety and without any changes. On request SIQ will give information about the validity of the licence.

Licence
for the use of SIQ certification mark

Number: **SI-SIQ 063/098**

Additional remarks:

The list of critical components of the product for which this licence is granted is included in the test report.

Places of manufacture, inspection file number:

1) GlobTek, Inc.
186 Veterans Drive Northvale, NJ 07647, USA
(Inspection file number: 15994-15994)

2) GlobTek (Suzhou) Co., Ltd.
Building 4, No. 76, Jinling East Road, Suzhou Industrial Park, JiangSu 215021, China
(Inspection file number: 15994-27052)

Ratings:

Input: 100-240 V a.c.; 50-60 Hz; 1,6 Amax.; Class I

Output: see table below:

Model	Standard model	Standard model output voltage (Vdc)	Standard model rated output current (A)	Y.Y scope	Output voltage range(V)	Max output power (W)
GTM21097-2003-Y.Y	GTM21097-2003	3.3	6	0.1-0.3 or blank	3-3,3	20
GTM21097-3005-Y.Y	GTM21097-3005	5	6	0.1-1.6 or blank	3,4-5	30
GTM21097-4509-Y.Y	GTM21097-4509	9	5	0.1-3.9 or blank	5,1-9	45
GTM21097-5012-Y.Y	GTM21097-5012	12	4,17	0.1-2.9 or blank	9,1-12	50
GTM21097-5015-Y.Y	GTM21097-5015	15	3,3	0.1-6.9 or blank	12,1-15	50
GTM21097-5018-Y.Y	GTM21097-5018	18	2,8	0.1-6.9 or blank	15,1-18	50
GTM21097-5024-Y.Y	GTM21097-5024	24	2,1	0.1-5.9 or blank	18,1-24	50
GTM21097-5048-Y.Y	GTM21097-5048	48	1,0	0.1-23.9 or blank	24,1-48	50

Limitation:

This Licence for the use of SIQ certification mark substitutes previously issued Licence for the use of SIQ certification mark No. SI-SIQ 063/084, issued on 2019-11-04, due to modification of the test report and standard update.

Date: 2024-09-30

Authorized signature: Matej Žontar