Ref. Certif. No.

SE-119632

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Drive Northvale NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory Note: *When more than one factory, please report on page 2*	☒ Additional Information on page 2
Ratings and principal characteristics	See page 2
Trademark / Brand (if any)	GlobTek, Inc., GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*41061-***
Additional information (if necessary may also be reported on page 2)	☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 IEC 60601-1-6:2010+A1+A2 IEC 60601-1-11:2015+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	2601B2316SHA-001, 2601B2316SHA-002, 2601B2316SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Date: 12 June, 2026

Signature:

Hyden Li

Factories

1. GlobTek, Inc.
186 Veterans Drive Northvale NJ 07647, USA

2. GlobTek (Suzhou) Co., Ltd
Building 4, No. 76 Jin Ling East Road, Suzhou Industrial Park, Suzhou, Jiang Su, 215021, China

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz or 50/60Hz, 0.6A;
Output: 5.0-30.0Vdc, Max.3.6A, Max. 18W

Model (3rd "*" has been filled)	Input value	Rated output voltage range	Max. rated output current	Max. rated output power	Transformer designations
GT*41061-*12*	100-240V~, 50-60Hz or 50/60Hz, 0.6A	5-12Vdc	3.6A	18W	TF001
GT*41061-*18*		12.1-18Vdc	1.49A	18W	TF002
GT*41061-*24*		18.1-24Vdc	1.0A	18W	TF003
GT*41061-*30*		24.1-30Vdc	0.75A	18W	TF004

Additional information

Explanation of model GT*41061-***:
The 1st "*" part can be 'M' or '-' or 'H' for market identification.
The 2nd "*" part denotes the rated output wattage designation, which can be "01" to "18", in step of 1 denote 1W.
The 3rd "*" part denotes the standard rated output voltage designation, which can be "12", "18", "24" or "30". These standard rated output voltage designations correspond to four isolated transformer models. Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.
The 4th "*" part can be "-0.1" to "-7.0" in step of 0.1 to denote voltage deviation or blank to indicate no voltage different. The result by subtracting the deviation value from the standard rated output voltage denotes the rated output voltage, with a range of 5 - 30 volts.

The group and national differences for the CENELEC countries and the national differences for Canada, USA, Switzerland, Korea, Japan have been checked.

The following applicable requirements have not been evaluated by Intertek:
-Biocompatibility according to ISO 10993
-EMC according to IEC 60601-1-2

Date: 12 June, 2026

Signature: