CERTIFICATE
FOR EUROPEAN PRODUCT SAFETY

Reference No. SE-S-2600602

Medical Power Supply

Type designation:	GT*41061-***
Certificate holder:	GlobTek, Inc. 186 Veterans Drive Northvale NJ 07647, USA
The product complies with the standard(s):	EN 60601-1:2006+A1+A12+A2+A13 EN 60601-1-6:2010+A1+A2 EN 60601-1-11:2015+A1
Date of expiry:	13 July, 2031
EU Directive information:	The product listed has been evaluated by a Third Party Body and deemed to meet the listed standard(s). This Certificate can be used in support of a claim in meeting relevant EU legislation and mandatory Conformity Marking. And in accordance with EU law, the claim is the sole obligation of the Manufacturer/ Importer.

Additional information in Appendix

Certification Body: Intertek Semko AB, Product Certification **Place:** Shanghai

Signed: **Date:** 13 July, 2026

Hyden Li

This Certificate is for the exclusive use of Intertek's client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Certificate. Only the Client is authorized to permit copying or distribution of this Certificate and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. This certificate is not issued under accreditation, by a non-approved body.

APPENDIX:

Reference No. SE-S-2600602

Technical Data:

Type designation	GT*41061-***
Rated power (W)	Max. 18
Frequency (Hz)	50-60 or 50/60
Input	100-240V~, 50-60Hz or 50/60Hz, 0.6A
Output	5.0-30.0Vdc, Max.3.6A, Max. 18W
Trademark	GlobTek, Inc. GlobTek, Inc.
Test Report	2601B2316SHA-004, 2601B2316SHA-005, 2601B2316SHA-006
Product information	GT*41061-*** The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety. The 2nd "*" part denotes the rated output wattage designation, which can be "01" to "18", in step of 1 denote 1W. The 3rd "*" part denotes the standard rated output voltage designation, which can be "12", "18", "24" or "30". These standard rated output voltage designations correspond to four isolated transformer models. Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil. The 4th "*" part can be "-0.1" to "-7.0" in step of 0.1 to denote voltage deviation or blank to indicate no voltage different. The result by subtracting the deviation value from the standard rated output voltage denotes the rated output voltage, with a range of 5 - 30 volts. (see appendix for details)

Appendix:

Model Ratings list

Model	Input value	Rated output voltage range	Max. rated output current	Max. rated output power	Transformer designations
GT*41061-*12*	100-240V~, 50-60Hz or 50/60Hz, 0.6A	5-12Vdc	3.6A	18W	TF001
GT*41061-*18*		12.1-18Vdc	1.49A	18W	TF002
GT*41061-*24*		18.1-24Vdc	1.0A	18W	TF003
GT*41061-*30*		24.1-30Vdc	0.75A	18W	TF004

This Certificate is for the exclusive use of Intertek's client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Certificate. Only the Client is authorized to permit copying or distribution of this Certificate and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. This certificate is not issued under accreditation, by a non-approved body.

APPENDIX:

Reference No. SE-S-2600602

Manufacturing Sites: GlobTek (Suzhou) Co., Ltd
Building 4, No. 76 Jin Ling East Road, Suzhou Industrial Park,
Suzhou, Jiang Su, 215021, China

The instruction for use shall be written in a language acceptable according to the national regulation in the country where the product is to be used.

If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

Certification Body: Intertek Semko AB, Product Certification **Place:** Shanghai

Signed: **Date:** 13 July, 2026

Hyden Li

This Certificate is for the exclusive use of Intertek's client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Certificate. Only the Client is authorized to permit copying or distribution of this Certificate and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. This certificate is not issued under accreditation, by a non-approved body.