



Ref. Certif. No.

SI-5833

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Direct plug-in power supply unit

Name and address of the applicant
Nom et adresse du demandeur

GlobTek, Inc.
186 Veterans Dr Northvale, NJ 07647, USA

Name and address of the manufacturer
Nom et adresse du fabricant

GlobTek, Inc.
186 Veterans Dr Northvale, NJ 07647, USA

Name and address of the factory
Nom et adresse de l'usine

1) GlobTek, Inc.
186 Veterans Dr Northvale, NJ 07647, USA

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2ème page

☒ Additional Information on page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

See page 2.

Trademark (if any)
Marque de fabrique (si elle existe)

GlobTek, Inc.

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.
Ref. De type

GTM41076-06xx ("xx" can be any rated output voltage
between 5 and 30 Vdc) See also page 2

Additional information (if necessary may also be reported
on page 2)
Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} page

See page 2

☒ Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1:2005 (Third Edition) + A1:2012

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

T221-0001/17 issued by laboratory SIQ Testing and
Certification GmbH

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



SIQ Ljubljana
Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

Date: 2017-02-17

Signature: Bojan Pečavar

Name and address of the factories:

1) Globtek Inc.,

186 Veterans Drive Northvale, NJ 07647, USA

2) GlobTek (Suzhou) Co., Ltd

Building 4, No. 76, Jinling East Road, Suzhou Industrial Park, Jiangsu 215021, China

3) Shenzhen Eng Electronics Co., Ltd.

Block B2, A4 first floor, A4 third-four Floor of the East, Nuclear Group Industrial District, Baishixia, Fuyun Town, Bao'an, Shenzhen China

Ratings and principal characteristics:

I/P: 100-240 V~; 50-60 Hz; 160-100 mA, Class II

O/P: See table below.

Model name	Output ratings (output DC voltage / output current)
GTM41076-0605	5,0 Vdc / 1,2 A
GTM41076-0606-0.9	5,1 Vdc / 1,176 A
GTM41076-0606	6,0 Vdc / 1,0 A
GTM41076-0607	7,0 Vdc / 0,85 A
GTM41076-0609-0.5	8,5 Vdc / 0,705 A
GTM41076-0609	9,0 Vdc / 0,66 A
GTM41076-0612	12,0 Vdc / 0,5 A
GTM41076-0615	15,0 Vdc / 0,4 A
GTM41076-0618	18,0 Vdc / 0,33 A
GTM41076-0620	20,0 Vdc / 0,3 A
GTM41076-0624	24,0 Vdc / 0,25 A
GTM41076-0630	30,0 Vdc / 0,16 A

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Certificate supersedes previously issued CB Test Certificate No. SI-3631, dated 2013-02-07, due to update of the test report.

Date: 2017-02-17

Signature: Bojan Pečavar

