

CERTIFICATE OF COMPLIANCE

Certificate Number 20131126-E341350
Report Reference E341350-A31-UL
Issue Date 2013-NOVEMBER-26

Issued to: GLOBTEK (HONG KONG) LTD
UNIT 1402, BENSON TOWER
74 HUNG TO RD, KWUN TONG
KOWLOON HONG KONG

**This is to certify that
representative samples of**



Power Supplies, Medical and Dental
Direct Plug-In Power Supply - GTM41076-06VV-X.X, were
VV can be any rated output voltage between 5 and 30 Vdc
and X.X is optional for specifying output voltage deviations,
X.X is to be subtracted from rated voltage.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1 and CAN/CSA-C22.2 No. 60601-1
(Medical Electrical Equipment - Part 1: General
Requirements for Basic Safety and Essential Performance)


Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

