		Ref. Certif. No.
IEC IECEE CB SCHEME		**SE-82078**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product Produit	Medical Power Supply
Name and address of the applicant Nom et adresse du demandeur	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647, United States
Name and address of the manufacturer Nom et adresse du fabricant	Same as above
Name and address of the factory Nom et adresse de l'usine Note: *When more than one factory, please report on page 2* Note: *Lorsque il y plus d'une usine, veuillez utiliser la 2ème*	See page 2
Ratings and principal characteristics Valeurs nominales et caractéristiques principales	Input: 100-240V~, 50-60Hz, 0.3A, Class II; Output: 5Vdc, Max. 1.2A
Trademark (if any) Marque de fabrique (si elle existe)	GlobTek, Inc.
Type of Manufacturer's Testing Laboratories used Type de programme du laboratoire d'essais constructeur	-
Model / Type Ref. Ref. De type	GT*41078-*05-USB
Additional information (if necessary may also be reported on page 2) Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2ème page)	See page 2
A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la	IEC 60601-1:2005+A1+C1+C2
As shown in the Test Report Ref. No. which forms part of this Certificate Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat	151202057SHA-001

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Intertek

Signature:

Date: 4 February 2016

Gary Hu

YUM

Factories
Les usines

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647, United States

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Additional information (if necessary)
Information complémentaire (si nécessaire)

Explanation of symbol "*" in type designation GT*41078-*05-USB:

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" part denotes the rated output wattage designation, which can be "01" to "06", with interval of 1.

Test performed on model GTM41078-0605-USB as worst case, and also performed on other output models for reference. Although ME equipment installation and use type are not specified, mechanical strength tests including push test, 1 m height drop test and mould stress relief test are performed considering the usage for adapter model.

Group differences for CENELEC countries and national difference for Canada, United States and Switzerland have been checked.

Date: 4 February 2016

Signature: