

CERTIFICATE OF COMPLIANCE

Certificate Number 20140228-E341350
Report Reference E341350-A33-UL
Issue Date 2014-FEBRUARY-28

Issued to: GLOBTEK (HONG KONG) LTD
UNIT 1402, BENSON TOWER, 74 HUNG TO RD,
KWUN TONG, KOWLOON HONG KONG

**This is to certify that
representative samples of** COMPONENT - POWER SUPPLIES, MEDICAL AND
DENTAL



See Next page for Models

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1 & CAN/CSA-C22.2 No. 60601-1 -
Medical Electrical Equipment - Part 1: General
Requirements for Basic Safety and Essential Performance


Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20140228-E341350
Report Reference E341350-A33-UL
Issue Date 2014-FEBRUARY-28

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Switching Power Supply,

GTM41134-***-**

The 1st “*” denote the rated output wattage designation, which can be “01” to “06”, with interval of 1.

The 2nd “*” denote the standard rated output voltage designation, which can be “03”, “04”, “06”, “12”, “15”, “18”, “24”, “36” or “48”.

The 3rd “*” is optional deviation, subtracted from standard output voltage, which can be “-0.1” to “-11.9” with interval of 0.1, or blank to indicate no voltage different.


The 2nd and 3rd “**” together denote the output voltage, with a range of 3.3 - 48 volts.

The 4th “*” can be ‘F’ to denote open frame model with connector which is fixing on the PCB or ‘T’ to denote open frame model with appliance inlet.

When the 4th “*” is ‘F’, the 5th “*” can be blank representing models with Class I connector or ‘W’ representing models with Class II connector.

When the 4th “*” is ‘T’, the 5th “*” can be ‘2’ representing models with Class II inlet or ‘3’ and “3A” representing model with two types of Class I inlets.

When -** is blank, presents direct plug-in model with interchangeable plug (Class II only)



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

