Ref. Certif. No.

SE-113905

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Drive Northvale NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory Note: *When more than one factory, please report on page 2*	☒ Additional Information on page 2
Ratings and principal characteristics	Input: 100-240V~, 50-60Hz, 1.5A; Output: 5-48Vdc, Max 6A, Max 60W
Trademark / Brand (if any)	GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*43007-*****
Additional information (if necessary may also be reported on page 2)	☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2
As shown in the Test Report Ref. No. which forms part of this Certificate	240200623SHA-001

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

Date: 10 May, 2024

intertek

Signature:

Gary Hu

Factories

1. GlobTek, Inc.
186 Veterans Drive Northvale NJ 07647, USA

2. GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Additional information

Explanation of model GT*43007-*****:
The 1st symbol "*" can be “M” or “-” or “H” for market identification and not related to safety.
The 2nd symbol "*" can be A, B, or C, is related to PCB size: A= 2”x3”, B=2”x4”, C=3”x5”. The different PCB sizes are only for installation purpose in end product with no safety spacing modification.
The 3rd symbol "*" can be “01” to “60”, with interval of 1, denote the rated output wattage designation.
The 4th symbol "*" can be “05”,”07”,”09”,”12”, “15”,”18”,”24”, “36”, or “48”, denote the standard rated output voltage designation.
The 5th symbol "*" can be “-0.1” to “-11.9” with interval of 0.1, or blank to indicate no voltage different, is optional deviation the subtracted from standard output voltage.
The 4th and 5th asterisks together denote the output voltage with a range of 5- 48 volts.
The 6th symbol "*" can be “-F” or “-FW”, “-F” represents Class I models and “-FW” represents Class II model.

The national differences for USA, Canada have been checked.

Only the requirements of IEC 60601-1, as documented in the accompanying test report, have been addressed.

The following applicable requirements have not been evaluated by Intertek:
-Biocompatibility according to ISO 10993
-EMC according to IEC 60601-1-2

Date: 10 May, 2024

Signature: