



Shenzhen SEM.Test Technology Co., Ltd.

1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,
Bao'an District, Shenzhen, P.R.C. (518101)

CERTIFICATE OF CONFORMITY

Certificate No.: SEM16060954

The following product has been tested by Shenzhen SEM.Test Technology Co., Ltd. with the listing standards and found in conformity with the **EC Directive 2014/30/EU**. It is possible to use CE marking to demonstrate the conformity with this **EMC Directive**.

Report No. : STR16068063E

Applicant : GlobTek, Inc
Address : 186 Veterans Dr. Northvale, NJ 07647 USA
Manufacturer : 1. GlobTek, Inc.
2. GlobTek (Suzhou) Co., Ltd
Address : 1.186 Veterans Dr. Northvale, NJ 07647 USA
2. Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Description of Product : Medical Power Supply/ ITE Power Supply
Model No. : Type of attachment
Trade Name : 

Test Standards : **EN 55032:2012+AC:2013**
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 60601-1-2:2015
EN 55024:2010

The referred test report(s) show that the product complies with the essential requirements in the above listed standards. The applicant is authorized to use this certificate in connection with the EC declaration of conformity according to Annex 1 of the Directive.



Test Laboratory

Jandy So

Supervisor

Date of Issue: Jun 27, 2016

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.: +86-755-33663309

E-mail: sem@semtest.com.cn

Website: www.semtest.com.cn



Shenzhen SEM.Test Technology Co., Ltd.

1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,
Bao'an District, Shenzhen, P.R.C. (518101)

ANNEX 1 TO ADDING MODEL

Certificate No. : SEM16060954

Model List

GT*43007-***** series

The 1st "*" can be "M" or "-" or "H" for market identification and not related to safety.

The 2nd * is A, B, or C and is related to PCB size: A= 2"x3", B=2"x4", C=3"x5"

The 3rd "*" denote the rated output wattage designation, which can be "01" to "60", with interval of 1.

The 4th "*"denote the standard rated output voltage designation, which can be "05", "07", "09", "12", "15", "18", "24", "36", "48".

The 5th "*" is optional deviation, subtracted from standard output voltage, which can be "-0,1" to "-11,9" with interval of 0,1 or blank to indicate no voltage different.

The 4th "*" and 5th "*" together denote the output voltage, with a range of 5- 48 volts.

The 6th "*" can be -F or -FW, -F means class I, -FW means class II

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.: +86-755-33663309

E-mail: sem@semtest.com.cn

Website: www.semtest.com.cn