



Ref. Certif. No.

SE-72765

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Medical Power Supply

Name and address of the applicant
Nom et adresse du demandeur

GlobTek Inc
186 Veterans Drive, Northvale NJ 07647,
UNITED STATES

Name and address of the manufacturer
Nom et adresse du fabricant

Same as above

Name and address of the factory
Nom et adresse de l'usine
Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

See page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Input: 100-240V~, 50-60Hz, 0.3A
Output: see page 2

Trademark (if any)
Marque de fabrique (si elle existe)

GlobTek

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

-

Model / Type Ref.
Ref. De type

GTM43033-***

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2^{ème} page)

See page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1:2005 and C1+C2

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

130300178SHA-001

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 26 April 2013

Intertek

Signature:

Gary Hu

**Factories
Les usines**

GlobTek Inc
186 Veterans Drive, Northvale NJ 07647,
UNITED STATES

GlobTek (Suzhou) Co., Ltd.
Building 4, No.76 JinLing East Road,
Suzhou Industrial Park, Suzhou JiangSu, 215021,
CHINA

**Additional information (if necessary)
Information complémentaire (si nécessaire)**

GTM43033-***:

The 1st "*" part denotes the rated output wattage designation, which can be "01" to "06", with interval of 1.

The 2nd "*" part denotes the standard rated output voltage designation, which can be "03", "04", "06", "12", "15", "18", "24", "36" or "48". These standard rated output voltage designations correspond to five isolated transformer models (See the appended table 8.10 for details). Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.

The 3rd "*" part is optional, which can be "-0.1" to "-11.9" with interval of 0.1 to denote voltage deviation or blank to indicate no voltage different. The result by subtracting the deviation value from the standard rated output voltage denotes the rated output voltage, with a range of 3 - 48 volts.

Model list

Model	Rated output voltage range	Max. rated output current	Max. rated output power
GTM43033-*03	3Vdc	2A	6W
GTM43033-*04*	3.1-4V	1.94A	6W
GTM43033-*06*	4.1-6V	1.46A	6W
GTM43033-*12*	6.1-12V	0.98A	6W
GTM43033-*15*	12.1-15V	0.50A	6W
GTM43033-*18*	15.1-18V	0.40A	6W
GTM43033-*24*	18.1-24V	0.33A	6W
GTM43033-*36*	24.1-36V	0.25A	6W
GTM43033-*48*	36.1-48V	0.16A	6W

Test performed on 3Vdc/2A output adapter model and 48Vdc/0.125A output adapter model as worst condition, and also performed on other output models for reference. Although ME equipment installation and use type are not specified, mechanical strength tests including push test, 1 m height drop test and mould stress relief test are performed considering the usage for adapter model.

The group and national differences for the CENELEC countries and Canada, USA, Switzerland have been checked.

Date: 26 April 2013

Signature: 