

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR
ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Drive, Northvale, NJ 07647, USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input:100-240 V~, 50-60 Hz, 0.45 A, Class II
Output:5.0-5.5 Vdc, Max. 3.2 A, Max. 16 W

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*46161-**-*

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1,
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate

191000380TWN-001, 191000382TWN-001

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Signature:

A handwritten signature in blue ink, appearing to read 'Ted Wu', is written over a horizontal line.

Ted Wu

Date: 18 March, 2020

Factories

- 1) GlobTek, Inc.
186 Veterans Drive, Northvale, NJ 07647, USA
- 2) GlobTek (Suzhou) Co., Ltd.
Building 4, No. 76, JinLing East Rd., Suzhou Industrial Park, Suzhou, JiangSu, 215021, China

Additional information (if necessary)

The national difference for Canada, USA and Japan has been checked.

The product has also been checked according to EN 60601-1:2006+A11:2011+A1:2013+A12 and EN 60601-1-11:2010.

Explanation for model GT*46161-**-* :

The 1st symbol "*" denotes "M" or "-" or "H" for market identification and not related to safety.

The 2nd symbol "*" denotes the rated output wattage designation, which can be "01" to "16", with interval of 1.

The 3rd symbol "*" denotes the standard rated output voltage designation, which can be "5.0" to "5.5" or "05" to "05.5" with interval of 0.1 Vdc.

The last symbol "*" denotes the types of output connector: -USB means Type 1 one USB type-A, -USB1A means Type 2 one USB type-A, -USB2A means Type 2 dual USB type-A and -USBC means Type 2 USB type-C.

The following applicable requirements have not been evaluated by Intertek:

- Biocompatibility according to ISO 10993
- EMC according to IEC 60601-1-2

Date: 18 March, 2020

Signature:

