



Shenzhen SEM.Test Technology Co., Ltd.
1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,
Bao'an District, Shenzhen, P.R.C. (518101)


CERTIFICATE OF CONFORMITY

Certificate No.: SEM18118330

The following product has been tested by Shenzhen SEM.Test Technology Co., Ltd. with the listing standards and found in conformity with the **EU Directive 2014/30/EU & 2014/53/EU Directive**. It is possible to use CE marking to demonstrate the conformity with this **EMC & RED Directive**.

Report No. : STR18118126E
Applicant : GlobTek, Inc.
Address : 186 Veterans Dr. Northvale, NJ 07647 USA

Manufacturer : 1.GlobTek, Inc.
2.GlobTek (Suzhou) Co., Ltd
Address : 1.186 Veterans Dr. Northvale, NJ 07647 USA
2.Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou,JiangSu 215021, China

Description of Product : ITE / Medical Power supply
Model No. : See Model List
Trade Name : 
Test Standards : EN 55032:2015/AC:2016-07
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 55035:2017
EN 60601-1-2:2015

The referred test report(s) show that the product complies with the essential requirements in the above listed standards. The applicant is authorized to use this certificate in connection with the EU declaration of conformity according to Annex 1 of the Directive.



Test Laboratory



Jandy So
Supervisor

Date of Issue: Nov. 20 2018



This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.: +86-755-33663309

E-mail: sem@semtest.com.cn

Website: www.semtest.com.cn



Shenzhen SEM.Test Technology Co., Ltd.

1/F, Building A, Hongwei Industrial Park, Liuxian 2nd Road,

Bao'an District, Shenzhen, P.R.C. (518101)

ANNEX 1 TO ADDING MODEL

Certificate No. : SEM18118330

Model List

GT*46161-**-USB

The 1st “*” can be “M” or “-” or “H” for market identification and not related to safety.

The 2nd “**” denote the rated output wattage designation, which can be “01” to “16”, with interval

The 3rd “*” denote the standard rated output voltage designation, which can be “5.0” to “5.5” or “05.5” with interval of 0,1.

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole production and other relevant Directives have to be observed.

Tel.: +86-755-33663308

Fax.: +86-755-33663309

E-mail: sem@semtest.com.cn

Website: www.semtest.com.cn