

# Certificate



for European Product Safety

Reference No. 1714973

## Medical Power supply

Type designation	GT*46161-**-USB
Certificate holder	<b>GlobTek, Inc.</b> 186 Veterans Dr. Northvale, NJ 07647, USA
The product complies with the standard(s)	EN 60601-1: 2006 + A1:2013 + A12:2014 EN 60601-1-11: 2015
Date of expiry	04 August 2022

Additional information in Appendix.

Certification Body Intertek Semko AB, Product Certification

Place Shanghai

Signed

Gary Hu

Date 04 August 2017  
Page 1 of 2



## Intertek Appendix

Reference No. 1714973

### Technical data

#### Type designation

GT\*46161-\*\*-USB

#### Input

100-240V~, 50-60Hz, 0.45A

#### Output

5.0-5.5Vdc, Max.16W

#### BRAND name:



### Product information

The 1<sup>st</sup> "\*" can be "M" or "-" or "H" for market identification and not related to safety.

The 2<sup>nd</sup> "\*" denote the rated output wattage designation, which can be "01" to "16", with interval of 1.

The 3<sup>rd</sup> "\*" denote the standard rated output voltage designation, which can be "5.0" to "5.5" or "05" to "05.5" with interval of 0.1.

### Manufacturing site(s)

GlobTek (Suzhou) Co., Ltd  
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou,  
JiangSu 215121,  
P.R.China

If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

Safety related parts of the user instructions shall be written in a language acceptable to the country where the product is to be used.

### Certification Body

Intertek Semko AB, Product Certification

Place Shanghai

### Signed

Gary Hu

Date 04 August 2017

*This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.*