



Ref. Certif. No.

SE-88242

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz, 1.0A
Output: 5-48Vdc, Max. 6A, Max. 40W

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT*46402-***

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 and A1
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate

170702098SHA-001, 170702098SHA-002

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 7 September 2017

Intertek

Signature:

Gary Hu



Ref. Certif. No.

SE-88242

Factories

- 1. GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA
- 2. GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, JinLing East Rd., Suzhou Industrial Park, Suzhou, JiangSu, 215021, China
- 3. Shenzhen ENG Electronics Co., Ltd.
Block B2, A4 first floor, A4 third-four Floor of the East, Nuclear Group Industrial District, Baishixia, Fuyun Town, Bao'an, Shenzhen, China

Additional information (if necessary)

Explanation of model GT*46402-***:

- The 1st symbol "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.
- The 2rd symbol "*" denotes the rated output wattage designation, with a maximum value of "40".
- The 3th symbol "*" denotes the standard rated output voltage designation, which can be "05" to "48" or "5.0" to "48.0".
- The last symbol "*" denote any six character = 0-9 or A-Z or () or [] or - or blank for marketing purposes.

Model	Rated output voltage range	Max. rated output current	Max. rated output power
GT*46402-***	5-48Vdc	6A	40W

Group differences for CENELEC countries and national difference for Canada, USA, Switzerland, Korea, Japan have been checked.

Date: 7 September 2017

Signature: 