Ref. Certif. No.

SE-114930

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Drive Northvale NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory Note: *When more than one factory, please report on page 2*	☒ Additional Information on page 2
Ratings and principal characteristics	Input: 100-240V~, 50-60Hz, 1.0A; Output: 5-48Vdc, Max. 6A, Max. 40W
Trademark / Brand (if any)	GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*46402-***
Additional information (if necessary may also be reported on page 2)	☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 IEC 60601-1-6:2010+A1+A2 IEC 60601-1-11:2015+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	2407B2585SHA-001, 2407B2585SHA-002, 2407B2585SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Date: 08 October, 2024

Signature:

Hyden Li

Factories

GlobTek, Inc.
186 Veterans Drive Northvale NJ 07647, USA

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Additional information

Explanation of model GT*46402-***:
The 1st symbol “*” can be “M” or “-“or “H” for market identification and not related to safety.
The 2nd symbol “*” can be “01” to “40”, with interval of 1W, denotes the rated output wattage designation.
The 3rd symbol “*” can be “05” to “48”, or “5.0” to “48.0” with interval of 0.1V, denotes the standard rated output voltage designation.
The last symbol “*” denotes any six character, which can be 0-9 or A-Z or () or [ ] or – or blank for marketing purposes.

The national differences for USA, Canada have been checked.

The following applicable requirements have not been evaluated by Intertek:
-Biocompatibility according to ISO 10993
-EMC according to IEC 60601-1-2

Date: 08 October, 2024

Signature: