IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647 USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory Note: *When more than one factory, please report on page 2*	See page 2
Ratings and principal characteristics	Input: 100-240V~, 50-60Hz, 0.3A; Output: 5-5.2VDC, Max.2A, Max 10W
Trademark (if any)	GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*86100-**-W2**
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 and A1 IEC 60601-1-11:2015
As shown in the Test Report Ref. No. which forms part of this Certificate	171201810SHA-001, 171201810SHA-002

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 7 August 2018

intertek

Signature: Gary Hu

Factories

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

GlobTek (Suzhou) Co., Ltd.
Building 4, No.76 JinLing East Road Suzhou Industrial Park, Suzhou, 215021, JiangSu,
P.R.China

Additional information (if necessary)

Group differences for CENELEC countries and national difference for Canada, USA, Switzerland, Korea and Japan have been checked.

Explanation of model GT*86100-**-W2**:
The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.
The 2nd "*" denotes the rated output wattage designation, which can be "01" to "10", with interval of 1.
The 3rd "*" denotes the rated output voltage designation, which can be "05","5.1","5.2" or "05","05.1","05.2".
The 4th "*" designates type of plug and can be E for European plug, U for British plug, blank for North American /Japan plug/Taiwan plug, C for China plug, I for India plug, A for Australia plug, K for Korea plug.
The 5th "*" can be "-USB" or blank, -USB denote the power supplies use USB port, when it is blank, denote the power supplies use DC output wires.

Date: 7 August 2018

Signature: