



Ref. Certif. No.

SE-102033

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647, United States of America
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	See page 2
Ratings and principal characteristics	Input: 100-240V~, 50-60Hz or 50/60Hz, 0.3A; Output: 5.95-24Vdc, Max.2.0A, Max.12W
Trademark (if any)	
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*86101-***-W2***
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 IEC 60601-1-6:2010+A1 IEC 60601-1-11:2015
As shown in the Test Report Ref. No. which forms part of this Certificate	200402485SHA-001, 200402485SHA-002, 200402485SHA-003

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Signature:

Gary Hu

Date: 14 August, 2020

Factories

GlobTek (Suzhou) Co., Ltd.
Building 4, No.76, JinLing East Road
Suzhou Industrial Park, Suzhou
215021 Jiangsu
China

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
United States of America

Additional information

Explanation of model GT*86101-***-W2***:

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" denotes the rated output wattage designation, which can be "01" to "12", with interval of 1.

The 3rd "*" denotes the standard rated output voltage designation, which can be "05.95", "07.5", "09", "12.0", "15.0", "24.0" or "5.95", "7.5", "9.0", "12", "15", "24".

The 4th "*" is optional deviation, subtracted from standard output voltage, which can be "-0.01" to "-8.9" with interval of 0.01, or blank to indicate no voltage different.

The 3rd "*" and 4th "*" together denote the output voltage, with a range of 5.95 -24Vdc.

The 5th "*" designates type of plug and can be E for European plug, U for British plug, blank for North American/Japan plug/Taiwan plug, C for China plug, A for Australia plug, K for Korea plug, I for India plug, AR for Argentina plug, BR for Brazil plug, SA or AF for South African plug.

The 6th "*" can be "-USB" or blank, -USB denote the power supplies use USB port. When it is blank, denote the power supplies use DC output wires.

The last "*" can be any six character 0 to 9 or A to Z, "()", "[]", "-" or blank for marketing purpose.

The group and national differences for the CENELEC countries, and the national differences for Canada, USA, Switzerland, Korea, Japan have been checked.

Date: 14 August, 2020

Signature: 