




Reference No. : WTX21X09095480E-1  
 Holder : GlobTek, Inc.  
 Address : 186 Veterans Dr. Northvale, NJ 07647 USA  
 Product : Medical/ITE Power Supply  
 Trade Name :   
 Model(s) : See Model List  
 Ratings : Refer to the test report

**Test Standards:**

- EN 55032:2015+A11:2020
- EN 55035:2017+A11:2020
- EN IEC 61000-3-2:2019
- EN 61000-3-3:2013+A1:2019

The above product has been tested by us with the listed standards and found in compliance with the European Electromagnetic Compatibility Directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

- EN 55032: Electromagnetic compatibility of multimedia equipment - Emission requirements
- EN 55035: Electromagnetic compatibility of multimedia equipment - Immunity requirements
- EN IEC 61000-3-2: Electromagnetic Compatibility (EMC) Part 3-2: Limits for harmonic current emissions (Equipment input current up to and including 16A per phase)
- EN 61000-3-3: Electromagnetic Compatibility (EMC) Part 3-3: Limitation of voltage supply systems for equipment with rated current  $\leq 16A$

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the conformity declaration, the CE marking as shown below can be affixed on the equipment.



The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn

# ANNEX 1 TO ADDING MODEL



Certificate No. : WTX21X09095480E-1  
 Model List

GTM96180-1811-2.0-T3, GTM91120-3024-T3A, GTM96300-3624-T2

GT\*96180-\*\*\*\*\*, GT\*96300-\*\*\*\*\*, GT\*91120-\*\*\*\*\*,

GT\*96180-\*\*\*\*\*,

The 1st “\*” part can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2nd “\*” denotes the rated output wattage designation, which can be “01” to “18”, with interval of 1.

The 3rd “\*” denotes the standard rated output voltage designation, which can be “07”, “11”, “17.9”, “30”, “38”, “48”, “54” or “56” ;

The 4th “\*” is optional deviation, subtracted from standard output voltage, which can be “-0.01” to “-12.0” with interval of 0.01, or blank to indicate no voltage different.

The 3rd “\*” and 4th “\*” together denote the output voltage, with a range of 5 - 56 volts.

The 5th “\*” = blank, it means wall plug in with interchangeable blade

=-T2 means desktop class II with C8 AC inlet  
 =-T2A means desktop class II with C18 AC inlet  
 =-T3 means desktop class I or class II with functional earth with C14 AC inlet  
 =-T3A means desktop class I or class II with functional earth with C6 AC inlet

The 6th “\*” = Blank or -AP or -PP or -SP

-AP (with baby board) stands for Active POE (full IEEE compliant)  
 -PP (no baby board) stands for Passive POE  
 -SP (no baby board) stands for Simple POE

The last \* denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

**Ratings**

When the 6th “\*” is blank:  
 GT\*96180-\*\*\*\*\*, Input: 100-240V~, 50-60Hz or 50/60Hz, 0.6A, Output: 5-48Vdc, Max. 3.6A, Max. 18W

When the 6th “\*” = -AP or -PP or - SP:  
 GT\*96180-\*\*\*\*\*, Input: 100-240V~, 50-60Hz or 50/60Hz, 0.6A, Output: 18-56Vdc, Max. 1.0A, Max. 18W

GT\*96300-\*\*\*\*\* and GT\*91120-\*\*\*\*\*

The 1st “\*” part can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2nd “\*” denotes the rated output wattage designation, which can be “01” to “36”, with interval of 1.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn

The 3rd “\*” denotes the standard rated output voltage designation, which can be “07.5” , “10.5” , “14.5” , “19.5” , “24” , “36” , “48” , “54” or “56” ;

The 4th “\*” is optional deviation, subtracted from standard output voltage, which can be “-0.01” to “-11.9” with interval of 0.01, or blank to indicate no voltage different.

The 3rd “\*” and 4th “\*” together denote the output voltage, with a range of 5 - 56 volts.

The 5th “\*” =-T2 means desktop class II with C8 AC inlet

=-T2A means desktop class II with C18 AC inlet

=-T3 means desktop class I or class II with functional earth with C14 AC inlet

=-T3A means desktop class I or class II with functional earth with C6 AC inlet

=-R2 means hybrid desktop housing class II with C8 AC inlet

=-R3A means hybrid desktop housing class I or class II with functional earth with C6 AC inlet

=-F means Open Frame class I or class II with functional earth

=-FW means Open Frame class II

=-P2 means Encapsulated class II

=-P3 means Encapsulated class I or class II with functional earth

The 6th “\*” = Blank or -AP or -PP or -SP

-AP (with baby board) stands for Active POE (full IEEE compliant)

-PP (no baby board) stands for Passive POE

-SP (no baby board) stands for Simple POE

The last \* denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

## Ratings

When the 6th “\*” is blank:

GT\*96300-\*\*\*\*\*, Input: 100-240V~, 50-60Hz or 50/60Hz, 1.0A, Output: 5-48Vdc, Max. 4.5A, Max. 36W

GT\*91120-\*\*\*\*\*, Input: 100-240V~, 50-60Hz or 50/60Hz, 1.5A, Output: 5-48Vdc, Max. 4A, Max. 30W

When the 6th “\*” = -AP or -PP or - SP:

GT\*96300-\*\*\*\*\*, Input: 100-240V~, 50-60Hz or 50/60Hz, 1.0A, Output: 18-56Vdc, Max. 2.0A, Max. 36W

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn