

UL TEST REPORT AND PROCEDURE

| | |
|------------------------------------|---|
| Standard: | UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety) |
| Certification Type: | Component Recognition |
| CCN: | QQHM2, QQHM8 (Power Supplies, Medical and Dental) |
| Product: | Switching Power Adapter |
| Model: | GT(M) 91120-WWVV-X.X-T2 series where "M" is for Medical, WW is the rated output wattage designation, with a maximum value of "30"; VV is the standard rated output voltage designation, with a maximum value of "48"; -X.X denotes the optional deviation, subtracted or added from standard output voltage in 0.1 volt increments or blank to indicate the no voltage different; |
| Rating: | Input: 100-240 Vac, 50-60 Hz, 1.5A Output: Refer to ID 07-06 |
| Applicant Name and Address: | GLOBTEK (HONG KONG) LTD UNIT 1402, BENSON TOWER 74 HUNG TO RD KWUN TONG KOWLOON HONG KONG |

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of Underwriters Laboratories Inc. ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

Issue Date: 2011-10-03

Page 2 of 15

Report Reference #

E341350-A3-UL

Prepared by: Steve Chiu
Underwriters Laboratories Inc.



Reviewed by: Calvin Tang
Underwriters Laboratories Inc.

