



Ref. Certif. No.  
**SE-108409**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	See page 2
Ratings and principal characteristics	See page 2
Trademark (if any)	
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*41134*****, GT*96060*****, GT-41134-0606-W2-TAB
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 IEC 60601-1-6:2010+A1+A2 IEC 60601-1-11:2015+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	220201762SHA-001, 220201762SHA-002, 220201762SHA-003

This CB Test Certificate is issued by the National Certification Body

**Intertek Semko AB**  
Torshamnsgatan 43  
Box 1103  
SE-164 22 Kista, Sweden

Date: 23 June, 2022

**intertek**

Signature:

  
Hyden Li

**Factories**

GlobTek, Inc.  
186 Veterans Dr. Northvale, NJ 07647, USA

GlobTek (Suzhou) Co., Ltd  
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

**Ratings and principal characteristics**

Input:

100-240V~, 50-60Hz, 0.3A or 0.6A for GT\*41134\*\*\*\*\* and GT\*96060\*\*\*\*\*;  
120V~, 60Hz, 0.3A for GT-41134-0606-W2-TAB

Output:

Model	voltage	Max. current	Max. power
GT*41134**03***, GT*96060**03***, GT*41134**03.3***, GT*96060**03.3***	3.3V	1.8A	6W
GT*41134**04***, GT*96060**04***	3.4-4V	1.76A	6W
GT*41134**06***, GT*96060**06***	4.1-6V	1.46A	6W
GT*41134**12***, GT*96060**12***	6.1-12V	0.98A	6W
GT*41134**15***, GT*96060**15***	12.1-15V	0.50A	6W
GT*41134**18***, GT*96060**18***	15.1-18V	0.40A	6W
GT*41134**24***, GT*96060**24***	18.1-24V	0.33A	6W
GT*41134**36***, GT*96060**36***	24.1-36V	0.25A	6W
GT*41134**48***, GT*96060**48***	36.1-48V	0.16A	6W
GT-41134-0606-W2-TAB	6V	1A	6W

**Additional information**

Explanation of models GT\*41134\*\*\*\*\*, GT\*96060\*\*\*\*\*:

The 1<sup>st</sup> symbol “\*\*” can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2<sup>nd</sup> symbol “\*\*” can be “-” or “CC”, “-” = Constant Voltage Model, CC = Constant Current Model.

The 3<sup>rd</sup> symbol “\*\*” denotes the rated output wattage designation, which can be “01” to “06”, with interval of 1.

The 4<sup>th</sup> symbol “\*\*” denotes the standard rated output voltage designation, which can be “03”, “03.3”, “04”, “06”, “12”, “15”, “18”, “24”, “36” or “48”.

The 5<sup>th</sup> symbol “\*\*” is optional deviation, subtracted from standard output voltage, which can be “-0.1” to “-11.9” with interval of 0.01, or blank to indicate no voltage different.

The 4<sup>th</sup> symbol “\*\*” and 5<sup>th</sup> symbol “\*\*” together denote the output voltage, with a range of 3.3 - 48 volts.

The 6<sup>th</sup> symbol “\*\*” =Blank means directly plug in model series,

= “-F” means Class I open frame model with connector which is fixing on the PCB,

= “-FW” means Class II open frame model with connector which is fixing on the PCB.

= “-FWT2” means open frame model with appliance inlet with Class II inlet C8 respectively,

= “-FT3A” means open frame model with appliance inlet with Class I inlet C6 respectively,

= “-FT3” means open frame model with appliance inlet with Class I inlet C14 respectively,

The last symbol “\*\*” denote any six character = 0-9 or A-Z or ([ ] or – or blank for marketing purposes.

Test performed on 3.3V, 5V, 9V and 48V output model as worst condition, and also performed on model GT-41134-0606-W2-TAB for reference. For open frame model, temperature rise, leakage current, protective earthing construction and mechanical strength shall be reevaluated on this component when installed in the end product.

Date: 23 June, 2022

Signature: 