




Reference No. : WTX23X06126282E
 Applicant : GlobTek, Inc.
 Address : 186 Veterans Dr. Northvale, NJ 07647 USA
 Manufacturer : 1: GlobTek, Inc. 2: GlobTek (Suzhou) Co., Ltd
 1: 186 Veterans Dr. Northvale, NJ 07647 USA
 Address : 2: Building 4, No. 76, Jin Ling East Rd., Suzhou
 Industrial Park, Suzhou, JiangSu 215021, China
 Product Name : Power Supply
 Trade Name : 
 Model(s) : See Model List
 Ratings : Refer to the test report

Test Standards:

EN 60601-1-2:2015+A1:2021

EN 60601-1-2: Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests



The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn

ANNEX 1 TO ADDING MODEL



Certificate No. : WTX23X06126282E

Model List

GT*41134*****, GT*96060*****

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" part can be "-" or "CC", "-" = Constant Voltage Model, CC = Constant Current Model.

The 3rd "*" denotes the rated output wattage designation, which can be "01" to "06", with interval of 1.

The 4th "*" denotes the standard rated output voltage designation, which can be "03", "04", "06", "12", "15", "18", "24", "36" or "48". These standard rated output voltage designations correspond to seven isolated transformer models.

The 5th "*" is optional deviation, subtracted from standard output voltage, which can be "-0.1" to "-11.9" with interval of 0.1, or blank to indicate no voltage different.

The 4th "*" and 5th "*" together denote the output voltage, with a range of 3.3 - 48 volts.

The 6th "*" = Blank means directly plug in model series,

= "-F" means Class I open frame model with connector which is fixing on the PCB,

= "-FW" means Class II open frame model with connector which is fixing on the PCB.

= "-FWT2" means open frame model with appliance inlet with Class II inlet C8 respectively,

= "-FT3A" means open frame model with appliance inlet with Class I inlet C6 respectively,

= "-FT3" means open frame model with appliance inlet with Class I inlet C14 respectively,

The last * denote any six character = 0-9 or A-Z or () [] or - or blank for marketing purposes.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn