

Reference No. : WTX24X07157877E002
Applicant : GlobTek, Inc.
Address : 186 Veterans Dr. Northvale, NJ 07647 USA
Manufacturer : GlobTek, Inc.
Address : 186 Veterans Dr. Northvale, NJ 07647 USA
Product Name : Power supply
Trade Name : 
Model(s) : See Model List
Ratings : Refer to the test report



Test Standards:

EN 60601-1-2:2015+A1:2021

EN 60601-1-2: Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests



Jandy So / GM
Date: 2024-08-02

The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Http://www.waltek.com.cn

Fax: +86-755-3366 3309

E-mail: sem@waltek.com.cn

ANNEX 1 TO ADDING MODEL

Certificate No. : WTX24X07157877E002

Model List

GTM961005P-*PD***

The 1st “*” =1 to 100, with interval of 1, denoting the rated output wattage designation from 1 W to 100 W.

The 2nd “*” = -USBCJ means USB Type-C jack in housing

= -USBCP means USB Type-C plug on fixed cord with strain-relief in housing

The 3rd “*” = -T2 means desktop class II with C8 AC inlet

= -T2A means desktop class II with C18 AC inlet

= -T3 means desktop class I with C14 AC inlet

= -T3A means desktop class I with C6 AC inlet

The last * denotes any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

Waltek Testing Group (Shenzhen) Co., Ltd.

Tel: +86-755-3366 3308

Fax: +86-755-3366 3309

Http://www.waltek.com.cn

E-mail: sem@waltek.com.cn