Ref. Certif. No.

DK-162252-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GLOBTEK (HONG KONG) LTD UNIT 1402, BENSON TOWER 74 HUNG TO RD KWUN TONG KOWLOON Hong Kong
Name and address of the manufacturer	GLOBTEK (HONG KONG) LTD UNIT 1402, BENSON TOWER 74 HUNG TO RD KWUN TONG KOWLOON Hong Kong
Name and address of the factory Note: *When more than one factory, please report on page 2*	GLOBTEK (SUZHOU) CO LTD BLDG 4, #76 JINLING EAST RD SUZHOU PARK SUZHOU JIANGSU 215021 China ☐ Additional Information on page 2
Ratings and principal characteristics	Input: 100-240Vac or 100-120Vac (for TP/TP3 models with NA standard plug), 50-60Hz, 1.5A Output: 5-54VDC, Max. 65W
Trademark / Brand (if any)	GlobTek, Inc.
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	GT*96600-**-** ☒ Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	☐ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-6:2010/AMD2:2020, IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E341350-D1012-1/A0/C0-CB issued on 2025-01-14

This CB Test Certificate is issued by the National Certification Body

UL Solutions

☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2025-01-21

Signature:

Thomas Wilson

Ref. Certif. No.

DK-162252-UL

Additional Model Detail(s):
GT*96600-**-**, The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety. The 2nd "*" denotes the rated output wattage designation, which can be "01" to "65", with interval of 1. The 3rd"*" denote the standard rated output voltage designation, which can be "05" to "54" or "5.0" to "54.0" in 0.1V increments. The 4th"*" can be "T2', "T2A", "T3", " T3A", "T2L", "T2AL", "T3L", " T3AL", "R2", "R3A", "P2", "P3", "TP", "TP3", "TW", "TW3", "T2" means desktop class II with C8 AC inlet, "T2A" means desktop class II with C18 AC inlet, "T3" means desktop class II with functional earth with C14 AC inlet, "T3A" means desktop class II with functional earth with C6 AC inlet, "T2L" means desktop class II with C8 AC inlet and housing with a DC jack, "T2AL" means desktop class II with C18 AC inlet and housing with a DC jack, "T3L" means desktop class II with functional earth with C14 AC inlet and housing with a DC jack, "T3AL" means desktop class II with functional earth with C6 AC inlet and housing with a DC jack, "R2" means hybrid desktop housing class II with C8 AC inlet, "R3A" means hybrid desktop housing class II with functional earth with C6 AC inlet, "P2" means Encapsulated class II, "P3" means Encapsulated class II with functional earth, "TP" means desktop class II with power supply cord with plug, "TP3" means desktop class II with functional earth with power supply cord with plug, "TW" means desktop class II with input wire without plug, "TW3" means desktop class II with functional earth with input wire without plug. The last * denote any six characters = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

Additional information (if necessary)

☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2025-01-21

Signature:

Thomas Wilson