



Ref. Certif. No.

SE-109578M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory	<input checked="" type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	See page 2
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*96605-G2*****
Additional information (if necessary may also be reported on page 2)	<input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1+A2 IEC 60601-1-6:2010+A1+A2 IEC 60601-1-11:2015+A1
As shown in the Test Report Ref. No. which forms part of this Certificate	220201764SHA-001, 220201764SHA-002, 220201764SHA-003, 220201764SHA-001 M1

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

Date: 21 September, 2023



Signature:

Hyden Li



Ref. Certif. No.

SE-109578M1

Factories

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647, USA

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Ratings and principal characteristics

Input: 1.5A, 100-240V~, 50-60Hz;
Output: 3.6-20VDC, Max.4.6A, Max. 60W

Model	Input	Output voltage (Vdc)	Max. output current (A)	Max. output power (W)
GT*96605-G2*****	100-240VAC, 50-60Hz, 1.5A	3.6V - 6.9V	4.6A	30.8W
		7.0V - 11.9V	4.4A	48.0W
		12.0V - 14.9V	4.0A	54.0W
		15.0V - 17.9V	3.6A	57.6W
		18.0V - 20.0V	3.2A	60.0W

Additional information

Explanation of model GT*96605-G2*****:

The 1st symbol “*” can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2nd symbol “*” can be A1, A2, A3 or any words for market use.

The 3rd symbol “*” denotes blank or the rated output wattage designation, which can be “01” to “60”, with interval of 0.1W.

The 4th symbol “*” =-T2 means desktop class II with C8 AC inlet.

=-T2A means desktop class II with C18 AC inlet.

=-T3 means desktop class I or class II with functional earth with C14 AC inlet.

=-T3A means desktop class I or class II with functional earth with C6 AC inlet.

=-T3F means desktop class I or class II with functional earth with C14 AC inlet with FLOATING OUTPUT.

=-T3AF means desktop class I or class II with functional earth with C6 AC inlet with FLOATING OUTPUT.

=-R2 means hybrid desktop housing class II with C8 AC inlet.

=-R3A means hybrid desktop housing class I or class II with functional earth with C6 AC inlet.

=-R3AF means hybrid desktop housing class I or class II with functional earth with C6 AC inlet with FLOATING OUTPUT.

The 5th symbol “*” can be blank or -RA. -RA denotes the Product with RIGHT ANGLE daughter board (no output cord); blank denotes the product with output cord.

The last symbol “*” denote any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

The whole series output will be any one voltage/current combinations (Power Profiles), between 3.6V and 20V.

The national differences for USA, Canada, Switzerland, Japan have been checked.

This certificate replaces CB certificate SE-109578, dated 22 November, 2022. A new certificate has been issued due to: Retested the 8.8.3 Dielectric strength on the prototype GTM96605-G2A1-T3.

Date: 21 September, 2023

Signature: