eurofins | E&E Certification Record

Listing#: E115461
Report #: RecogC 129956 c1 139539
Original Certification Date: April 12, 2024
Revised Certification Date: November 11, 2025

This Certification is issued to:
GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647 USA

Stating that the product(s):
Medical Power Supply
Series GT*9X1001P-*****

Product Rating(s):
- Input: 100-240V~,50-60Hz or 50/60Hz, 1.5A; 115Vac, 400Hz, 1.5A
- Output: 12.0-54.0V , Max.8.33A, Max. 100.0W

Achieved Certification to the following standard(s):
ANSI/ AAMI ES60601-1:2005, ES60601-1:2005/AMD1 1:2012, ES60601-1:2005/AMD 2:2021
CAN/CSA-C22.2 No. 60601-1:14 + A2:22 (R2022)
Medical electrical equipment— Part 1: General requirements for basic safety and essential performance
IEC 60601-1-6 Edition 3.2 2020-07 /
CAN/CSA-C22.2 NO. 60601-1-6:11 + A1:15 + A2:21 (R2021)
Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
ANSI/ AAMI HA60601-1-11:2015 [Including AMD1: 2021]
CSA C22.2 NO. 60601-1-11:15 (R2020) + A1:21
Medical Electrical Equipment -- Part 1-11: General requirements for basic safety and essential performance -- Collateral Standard: Requirements for medical electrical equipment and medical electrical equipment and medical electrical systems used in the home healthcare environment
ANSI AAMI IEC 60601-1-12:2016 [Including AMD 1:2021]
CSA C22.2 NO. 60601-1-12:15 (R2020) +A1:21
Medical electrical equipment - Part 1-12: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency

Hon Keung Ip
Certification Officer,
Eurofins Electrical and Electronic Testing North America, LLC

All changes proposed in the previously identified product that affects the above information must be submitted to Eurofins for evaluation prior to implementation to assure continued MET Recognition status. The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by Eurofins E&E NA and that all responsibilities are being fulfilled as specified in the Applicants' Responsibility section of the Certification Report. The Applicant named above has been authorized by Eurofins E&E NA to represent the product(s) listed in this record as "MET Recognized" and to mark this/these product(s) according to the terms and conditions of the Eurofins E&E NA Applicant Contract, Listing Reports, and the applicable agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in this Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.

Eurofins/MET is the first Nationally Recognized Testing Laboratory (NRTL) in the United States for testing and certifying products to UL safety standards.