

CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150624
Issue Date 2015-OCTOBER-20

Issued to: GLOBTEK INC
186 VETERANS DR
NORTHVALE, NJ 07647
United States

This is to certify that representative samples of POWER SUPPLIES, MEDICAL AND DENTAL - COMPONENT
For models refer to Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety: ANSI/AAMI ES 60601-1: 2005 / CAN/CSA C22.2, No. 60601-1:08 - Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance.

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150624
Issue Date 2015-OCTOBER-20

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Models - GTMG3069 series, M can be "M" or "-" for market purpose
GTMG3069-0505-PP, GTMG3069-0512-PP, GTMG3069-0515-PP, GTMG3069-0524-PP.
GTMG3074 series, M can be "M" or "-" for market purpose
GTMG3074-0505-FW, GTMG3074-0512-FW, GTMG3074-0515-FW, GTMG3074-0524-FW.
GTMG3070 series, M can be "M" or "-" for market purpose
GTMG3070-1005-PP, GTMG3070-1012-PP, GTMG3070-1015-PP, GTMG3070-1024-PP.
GTMG3075 series, M can be "M" or "-" for market purpose
GTMG3075-1005-FW, GTMG3075-1012-FW, GTMG3075-1015-FW, GTMG3075-1024-FW.
GTMG3072 series, M can be "M" or "-" for market purpose.
GTMG3072-2005-PP, GTMG3072-2012-PP, GTMG3072-2015-PP, GTMG3072-2024-PP.
GTMG3077 series, M can be "M" or "-" for market purpose
GTMG3077-2005-FW, GTMG3077-2012-FW, GTMG3077-2015-FW, GTMG3077-2024-FW.
GTMG3073 series, M can be "M" or "-" for market purpose.
GTMG3073P120012-C, GTMG3073P120015-C, GTMG3073P120024-C, GTMG3073P120036-
3.0-C, GTMG3073P120036-C, GTMG3073P120048-C, GTMG3073P120060-C



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150625
Issue Date 2015-OCTOBER-20

Issued to: GLOBTEK INC
186 VETERANS DR
NORTHVALE , NJ 07647
United States

This is to certify that representative samples of POWER SUPPLIES, MEDICAL AND DENTAL - COMPONENT

For models refer to Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety: UL 60601-1 and CAN/CSA-C22.2 No. 601.1-M90, 2005 – (Medical Electrical Equipment - Part 1: General Requirements for Safety)

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>




CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150625
Issue Date 2015-OCTOBER-20

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Models - GTMG3069 series, M can be "M" or "-" for market purpose
GTMG3069-0505-PP, GTMG3069-0512-PP, GTMG3069-0515-PP, GTMG3069-0524-PP
GTMG3074 series, M can be "M" or "-" for market purpose
GTMG3074-0505-FW, GTMG3074-0512-FW, GTMG3074-0515-FW, GTMG3074-0524-FW
GTMG3070 series, M can be "M" or "-" for market purpose
GTMG3070-1005-PP, GTMG3070-1012-PP, GTMG3070-1015-PP, GTMG3070-1024-PP
GTMG3075 series, M can be "M" or "-" for market purpose
GTMG3075-1005-FW, GTMG3075-1012-FW, GTMG3075-1015-FW, GTMG3075-1024-FW
GTMG3071 series, M can be "M" or "-" for market purpose
GTMG3071-1505-PP, GTMG3071-1512-PP, GTMG3071-1515-PP, GTMG3071-1524-PP
GTMG3076 series, M can be "M" or "-" for market purpose
GTMG3076-1505-FW, GTMG3076-1512-FW, GTMG3076-1515-FW, GTMG3076-1524-FW
GTMG3072 series, M can be "M" or "-" for market purpose
GTMG3072-2005-PP, GTMG3072-2012-PP, GTMG3072-2015-PP, GTMG3072-2024-PP
GTMG3077 series, M can be "M" or "-" for market purpose
GTMG3077-2005-FW, GTMG3077-2012-FW, GTMG3077-2015-FW, GTMG3077-2024-FW
GTMG3081 series, M can be "M" or "-" for market purpose
GTMG3081-400512-FW, GTMG3081-440524-FW, GTMG3081-4512-FW,
GTMG3081-4515-FW, GTMG3081-4524-FW, GTMG3081-4548-FW,
GTMG3082 series, M can be "M" or "-" for market purpose
GTMG3082-490512-FW, GTMG3082-540524-FW, GTMG3082-6012-FW,
GTMG3082-6015-FW, GTMG3082-6024-FW, GTMG3082-6048-FW



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150626
Issue Date 2015-OCTOBER-20

Issued to: GLOBTEK INC
186 VETERANS DR
NORTHVALE , NJ 07647
United States

This is to certify that representative samples of POWER SUPPLIES, MEDICAL AND DENTAL - COMPONENT

For models refer to Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1 and CAN/CSA-C22.2 No. 60601-1 - (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance)

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20151020-E172861
Report Reference E172861-20150626
Issue Date 2015-OCTOBER-20


This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

GTMG3083 series, M can be "M" or "-" for market purpose
GTMG3083P15012-FW, GTMG3083P15015-FW, GTMG3083P15024-FW,
GTMG3083P15048-FW

GTMG3080 series, M can be "M" or "-" for market purpose
GTMG3080P25012-FW, GTMG3080P25015-FW, GTMG3080P25124-FW,
GTMG3080P25248-FW

GTMG3078 series, M can be "M" or "-" for market purpose
GTMG3078P35012-FW, GTMG3078P35015-FW, GTMG3078P35124-FW,
GTMG3078P35148-FW

GTMG3079 series, M can be "M" or "-" for market purpose
GTMG3079P45012-F, GTMG3079P45015-F, GTMG3079P45024-F,
GTMG3079P45048-F



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

