



Ref. Certif. No.

SE-81897

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Medical Power supply

Name and address of the applicant
Nom et adresse du demandeur

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
USA

Name and address of the manufacturer
Nom et adresse du fabricant

Same as applicant

Name and address of the factory
Nom et adresse de l'usine
Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

See page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Input:
GT*41134*****, GT*96060*****: 100-240V~, 50-60Hz, 0.3A or 0.6A; GT-41134-0606-W2-TAB: 120V~, 60Hz, 0.3A

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

-

Model / Type Ref.
Ref. De type

GT*41134*****, GT*96060*****, GT-41134-0606-W2-TAB

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2^{ème} page)

See page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1:2005+A1
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

151100933SHA-001, 151100933SHA-002

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Intertek

Signature:

Gary Hu

Date: 21 January 2016

**Factories
Les usines**

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647,
USA

GlobTek (Suzhou) Co., Ltd.
Building 4, No.76 JinLing East Road, Suzhou Industrial Park, Suzhou JiangSu, 215021,
China

**Additional information (if necessary)
Information complémentaire (si nécessaire)**

GT*41134***** and GT*96060*****:

The first "*" part can be 'M' or 'L' or 'H' for market identification and not related to safety.

The second "*" part can be "-" or "CC", "-" = Constant Voltage Model, CC = Constant Current Model.

The third "*" denotes the rated output wattage designation, which can be "01" to "06", with interval of 1.

The fourth "*" denotes the standard rated output voltage designation, which can be "03", "04", "06", "12", "15", "18", "24", "36" or "48". These standard rated output voltage designations correspond to seven isolated transformer models (See the appended table 8.10 for details). Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.

The fifth "*" is optional deviation, subtracted from standard output voltage, which can be "-0.1" to "-11.9" with interval of 0.1, or blank to indicate no voltage different.

The fourth "*" and fifth "*" together denote the output voltage, with a range of 3.3 - 48 volts.

The sixth "*" = Blank means directly plug in model series,

= "-F" means Class I open frame model with connector which is fixing on the PCB,

= "-FW" means Class II open frame model with connector which is fixing on the PCB.

= "-FWT2" means open frame model with appliance inlet with Class II inlet C8 respectively,

= "-FT3A" means open frame model with appliance inlet with Class I inlet C6 respectively,

= "-FT3" means open frame model with appliance inlet with Class I inlet C14 respectively,

The seventh denote any six character = 0-9 or A-Z or () [] or - or blank for marketing purposes.

Test performed on 3.3V, 5V, 9V and 48V output model as worst condition, and also performed on model GT-41134-0606-W2-TAB for reference. For open frame model, temperature rise, leakage current, protective earthing construction and mechanical strength shall be reevaluated on this component when installed in the end product.

Group differences for CENELEC countries and national difference for Canada, USA, Switzerland, Korea and Japan have been checked.

Date: 21 January 2016

Signature: 

Model list

| Model | voltage | Max. current | Max. power |
|------------------------------------|----------|--------------|------------|
| GT*41134**03*** GT*96060**03*** | 3.3V | 1.8A | 6W |
| GT*41134**04*** GT*96060**04*** | 3.4-4V | 1.76A | 6W |
| GT*41134**06*** GT*96060**06*** | 4.1-6V | 1.46A | 6W |
| GT*41134**12*** GT*96060**12*** | 6.1-12V | 0.98A | 6W |
| GT*41134**15*** GT*96060**15*** | 12.1-15V | 0.50A | 6W |
| GT*41134**18*** GT*96060**18*** | 15.1-18V | 0.40A | 6W |
| GT*41134**24*** GT*96060**24*** | 18.1-24V | 0.33A | 6W |
| GT*41134**36*** GT*96060**36*** | 24.1-36V | 0.25A | 6W |
| GT*41134**48*** GT*96060**48*** | 36.1-48V | 0.16A | 6W |
| GT-41134-0606-W2-TAB | 6V | 1A | 6W |

Date: 21 January 2016

Signature: 