

No. P00102358


Order No. 200043151

Applicant GlobTek, Inc.
186 Veterans Dr.
Northvale
NJ 07647
USA

Manufacturer GlobTek, Inc.
186 Veterans Dr.
Northvale
NJ 07647
USA

Factory See page two of test report.

Group 61 91000

Model/type  **GlobTek, Inc.** GT500148-12

Data Input: 1.0A, 100-240VAC, 50-60Hz
Output: 12VDC, 48VA,

Other specification Class I equipment,

The above product is certified according to the following standard(s)

Safety std.: EN 60601-1 (1990) + A1 (1993) + A2 (1995) + A12 (1993) + A13 (1996)+ Corrigenda (July 1994)

EMC std.: EN 60601-1-2 (1993)

Validity The certificate is valid until 1. January 2011, provided that all signed certification conditions are complied with, and that possible changes to the product are notified to Nemko for acceptance prior to implementation. The validity time may be reduced in case new standards are made applicable. The certificate also applies as licence for use of Nemko's name and certification mark.

Variants This Certificate also covers variants with Position No. from 001 to 002
See next page(s)

Additional Information:


Date of issue 15 December 2000



signature

Arild Hansgård

Head of section



signature

Vegard Andersen