

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR
ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Multi-Unit Charger

Name and address of the applicant

GlobTek, Inc.
186 Veterans Dr. Northvale,
NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Class I, IP20,
Input: 100-240V~, 50-60Hz, 1.5A;
Output: 9-12.6VDC, 4A
AC inlet: Max. 10A; AC outlet: Max. 6.3A, Max. 1.5kW

Trademark (if any)

Dräger

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

GT-93600SHG3050

Additional information (if necessary may also be
reported on page 2)

See page 2

A sample of the product was tested and found
to be in conformity with

IEC 60335-1:2010+A1+A2
IEC 60335-2-29:2016

As shown in the Test Report Ref. No. which forms part
of this Certificate

171201041SHA-001,
171201041SHA-001 M1

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 26 October 2018

intertek

Signature:

Hyden Li



Ref. Certif. No.

SE-90070M1

Factories

GlobTek, Inc.
186 Veterans Dr. Northvale,
NJ 07647 USA

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76 JinLing East Road, Suzhou
Industrial Park, Suzhou, JiangSu, 215021, China

Additional information (if necessary)

Group differences for CENELEC countries and national differences for Australia/New Zealand are considered.

This certificate replaces CB certificate SE-90070, dated 5 Mar 2018.

A new certificate has been issued due to:

1. Changed the IP class from "IP30" to "IP20".
2. Changed the appliance outlet rating from "Max 8.5A" to "6.3A MAX, 1.5kW MAX".
3. Added a new supplier for PCB.
4. Added new supplies for new added fuse used near appliance outlet.
5. Added a new supplier for magnet wire.
6. Added new suppliers for enclosure.
7. Added national difference for Australian and New Zealand.
8. Added new photos for new added fuse used near appliance outlet.

Date: 26 October 2018

Signature: