

# CERTIFICATE

## FOR EUROPEAN PRODUCT SAFETY

Reference No. SE-S-2001153

### Medical Power supply

**Type designation:** GT\*41080\*\*\*\*\*


**Certificate holder:** GlobTek, Inc.  
186 Veterans Drive  
NJ 07647 Northvale  
United States

**The product complies with the standard(s):** EN 60601-1:2006+A1+A12

**Date of expiry:** 06 March, 2025

Additional information in Appendix

**Certification Body:** Intertek Semko AB, Product Certification      **Place:** Kista - Stockholm

**Signed:**       **Date:** 06 March, 2020

\_\_\_\_\_  
Fredrik Wennersten

## APPENDIX:

Reference No. SE-S-2001153

### Technical Data:

*Type designation* GT\*41080\*\*\*\*\*  
*Input* 100-240V~, 50-60Hz, 0.6A  
*Output* 5-48V, Max.3.6A, Max.18W  
*Trademark* GlobTek  
*Test Report* 200200842SHA-001  
*Product information* GT\*41080\*\*\*\*\* The 1st "\*" part can be 'M' or '-' or 'H' for market identification and not related to safety.  
The 2nd "\*" part can be "-" or "CC", "-" = Constant Voltage Model, CC = Constant Current Model.  
The 3rd "\*" part denotes the rated output wattage designation, which can be "01" to "18", with interval of 1.  
The 4rd "\*" part denotes the standard rated output voltage designation, which can be "07", "11", "17.9", "30", "38" and "48". Each standard rated output voltage designation corresponds to a transformer model. Each transformer model is identical in insulation construction including clearance and creepage except number of turns per coil.  
The 5th "\*" part is optional, which can be "-0.1" to "-12.0" with interval of 0.1 to denote voltage deviation or blank to indicate no voltage different. The result by subtracting the deviation value from the standard rated output voltage denotes the rated output voltage, with a range of 5 - 48 volts.  
The 6th "\*" part is also optional, which can be "-F" to denote open frame power supply model series or blank to denote direct plug-in power adapter model series.

### Manufacturing Sites:

GlobTek (Suzhou) Co., Ltd.  
Building 4, No.76, JinLing East Road, Suzhou Industrial Park, Suzhou,  
215021 Jiangsu  
China

Safety related parts of the user instructions shall be written in a language acceptable to the country where the product is to be used.


If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

### Certification Body:

Intertek Semko AB, Product Certification

**Place:** Kista - Stockholm

### Signed:



**Date:** 06 March, 2020

Fredrik Wennersten