



Ref. Certif. No.

**SE-84425**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product

Medical Power Supply

Name and address of the applicant

GlobTek, Inc.  
186 Veterans Dr. Northvale, NJ 07647 USA

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz, 0.6A / 1.0A / 1.5A;  
Output: 5-48VDC, Max 36W

Trademark (if any)

GlobTek

Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT\*\*-\*\*\*\*\*

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 and A1  
IEC 60601-1-11:2015

As shown in the Test Report Ref. No. which forms part of this Certificate

160100305SHA-001  
160100305SHA-002

This CB Test Certificate is issued by the National Certification Body

**Intertek Semko AB**  
Box 1103  
SE-164 22 Kista, Sweden  
Int +46 8 750 00 00

Date: 15 September 2016

**Intertek**

Signature:

Gary Hu

**Factories**

GlobTek, Inc.  
186 Veterans Dr. Northvale, NJ 07647 USA

GlobTek (Suzhou) Co., Ltd  
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

**Additional information (if necessary)**

Group differences for CENELEC countries and national difference for Canada, USA and Switzerland have been checked.

**Model Differences:**

GT\*\*\_\*\*\*\*\*

The 1st "\*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "\*" can be 96180 or 96300 or 91120 or 91128 for market identification

The 3rd "\*" denotes the rated output wattage designation, which can be "01" to "36", with interval of 1.

The 4th "\*" denotes the standard rated output voltage designation, when the 2nd "\*" = 96180 which can be "07", "11", "17.9", "30", "38" or "48"; when the 2nd "\*" = 96300 or 91120 which can be "07.5", "10.5", "14.5", "19.5", "24", "36" or "48".

The 5th "\*" is optional deviation, subtracted from standard output voltage, which can be "-0.01" to "-12.0" with interval of 0.01, or blank to indicate no voltage different.

The 4th "\*" and 5th "\*" together denote the output voltage, with a range of 5 - 48 volts.

The 6th "\*" = blank, it means wall plug in with interchangeable blade

--T2 means desktop class II with C8 AC inlet

--T2A means desktop class II with C18 AC inlet

--T3 means desktop class I with C14 AC inlet

--T3A means desktop class I with C6 AC inlet

--R2 means hybrid desktop housing class II with C8 AC inlet

--R3A means hybrid desktop housing class I with C6 AC inlet

--F means Open Frame class I

--FW means Open Frame class II

--P2 means Encapsulated class II

--P3 means Encapsulated class I

When the 2nd "\*" = 91128,

the model will be GTM91128LI1CEL Output: 4.2V, 1000mA;

or Model GTM91128LI2CEL Output: 8.4V, 1000mA;

or Model GTM91128LI3CEL Output: 12.6V, 1000mA;

The last \* denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

**Ratings:**

When 2nd "\*" = 96180, Input: 100-240V~, 50-60Hz, 0.6A Output: 5-48Vdc

When 2nd "\*" = 96300, 91120 or 91128, Input: 100-240V~, 50-60Hz, 1.5A or 1.0A Output: 5-48Vdc

Date: 15 September 2016

Signature: 