



Ref. Certif. No.
SE-101484M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	GlobTek, Inc. 186 Veterans Dr. Northvale, NJ 07647, USA
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	See page 2
Ratings and principal characteristics	Input: 100-240V~, 50-60Hz, 1.0A Output: 5-56VDC, Max 36W
Trademark (if any)	
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	GT*96300-*****
Additional information (if necessary may also be reported on page 2)	See page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005+A1 IEC 60601-1-11:2015
As shown in the Test Report Ref. No. which forms part of this Certificate	191201633SHA-001, 191201633SHA-002, 191201633SHA-001 M1, 191201633SHA-002 M1

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek
Signature: 
Gary Hu

Date: 30 September, 2020

Factories

GlobTek, Inc.
186 Veterans Dr. Northvale, NJ 07647, USA

GlobTek (Suzhou) Co., Ltd
Building 4, No. 76, Jin Ling East Rd., Suzhou Industrial Park, Suzhou, JiangSu 215021, China

Additional information

GT*96300-*****:

The 1st "*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "*" denotes the rated output wattage designation, which can be "01" to "36", with interval of 1.

The 3rd "*" denotes the standard rated output voltage designation, which can be "07.5", "10.5", "14.5", "19.5", "24", "36", "48", "54" or "56".

The 4th "*" is optional deviation, subtracted from standard output voltage, which can be "-0.01" to "-11.9" with interval of 0.01, or blank to indicate no voltage different.

The 3rd "*" and 4th "*" together denote the output voltage, with a range of 5 - 56 volts.

The 5th "*" =-T2 means desktop class II with C8 AC inlet

=-T2A means desktop class II with C18 AC inlet

=-T3 means desktop class I or class II with functional earth with C14 AC inlet

=-T3A means desktop class I or class II with functional earth with C6 AC inlet

=-R2 means hybrid desktop housing class II with C8 AC inlet

=-R3A means hybrid desktop housing class I or class II with functional earth with C6 AC inlet

The 6th "*" = Blank or -AP or -PP or -SP

-AP (with baby board) stands for Active POE (full IEEE compliant)

-PP (no baby board) stands for Passive POE

-SP (no baby board) stands for Simple POE

The last "*" can be any six character consist 0 to 9 or A to Z or () or - or blank for marketing purpose

The group and national differences for the CENELEC countries and Canada, USA, Switzerland have been checked.

This certificate replaces CB certificate SE-101484, dated 22 June 2020. A new certificate has been issued due to:
Added construction diagram of optional output connector as required by applicant.

Date: 30 September, 2020

Signature: 