

Reference No. : WTX21X09095460E-1  
 Holder : GlobTek (Suzhou) Co., Ltd  
 Address : Building 4, No. 76, Jin Ling East Rd., Suzhou  
 Industrial Park, Suzhou, JiangSu 215021, China  
 Product : ICT/ITE/Medical Power Supply  
 Trade Name :   
 Model(s) : See Model List  
 Ratings : Refer to the test report



## Test Standards:

EN 55032:2015+A11:2020  
 EN 55035:2017+A11:2020  
 EN IEC 61000-3-2:2019  
 EN 61000-3-3:2013+A1:2019

The above product has been tested by us with the listed standards and found in compliance with the European Electromagnetic Compatibility Directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

**EN 55032: Electromagnetic compatibility of multimedia equipment - Emission requirements**  
**EN 55035: Electromagnetic compatibility of multimedia equipment - Immunity requirements**  
**EN IEC 61000-3-2: Electromagnetic Compatibility (EMC) Part 3-2: Limits for harmonic current emissions (Equipment input current up to and including 16A per phase)**  
**EN 61000-3-3: Electromagnetic Compatibility (EMC) Part 3-3: Limitation of voltage supply systems for equipment with rated current  $\leq 16A$**

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the above mentioned EU Directive. Other relevant Directives have to be observed.

After preparation of the necessary technical documentation as well as the conformity declaration, the CE marking as shown below can be affixed on the equipment.



The statement is based on a single evaluation of the sample of above mentioned product(s). It does not imply an assessment of the whole production.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

Tel: +86-755-3366 3308  
 Http://www.waltek.com.cn

Fax: +86-755-3366 3309  
 E-mail: sem@waltek.com.cn

## ANNEX 1 TO ADDING MODEL

**Certificate No. : WTX21X09095460E-1**

### Model List

#### **GT\*96600-\*\*\*\***

The 1st “\*” part can be ‘M’ or ‘-’ or ‘H’ for market identification and not related to safety.

The 2nd “\*” denotes the rated output wattage designation, which can be “01” to “65”, with interval of 1.

The 3rd “\*” denote the standard rated output voltage designation, which can be “05” to “54” or “5.0” to “54.0” in 0.1V increments.

The 4th “\*” =-T2 means desktop class II with C8 AC inlet

=-T2A means desktop class II with C18 AC inlet

=-T3 means desktop class I or class II with functional earth with C14 AC inlet

=-T3A means desktop class I or class II with functional earth with C6 AC inlet

=-T2L means desktop class II with C8 AC inlet and housing with a DC jack

=-T2AL means desktop class II with C18 AC inlet and housing with a DC jack

=-T3L means desktop class I or class II with functional earth with C14 AC inlet and housing with a DC jack

=-T3AL means desktop class I or class II with functional earth with C6 AC inlet and housing with a DC jack

=-R2 means hybrid desktop housing class II with C8 AC inlet

=-R3A means hybrid desktop housing class I or class II with functional earth with C6 AC inlet

=-F means Open Frame class I or class II with functional earth

=-FW means Open Frame class II

=-P2 means Encapsulated class II

=-P3 means Encapsulated class I or class II with functional earth

=-TP means desktop class II with power supply cord with plug

=-TP3 means desktop class I or class II with functional earth with power supply cord with plug

=-TW means desktop class II with input wire without plug

=-TW3 means desktop class I or class II with functional earth with input wire without plug

The last “\*” denote any six character = 0-9 or A-Z or ()[] or – or blank for marketing purposes.

**Waltek Testing Group (Shenzhen) Co., Ltd.**

**Tel: +86-755-3366 3308**

**Http://www.waltek.com.cn**

**Fax: +86-755-3366 3309**

**E-mail: sem@waltek.com.cn**