



Ref. Certif. No.

**SE-91110**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product

Medical Power Supply

Name and address of the applicant

Globtek, Inc.  
186 Veterans Drive Northvale, NJ 07647 U.S.A.

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

See page 2

Ratings and principal characteristics

Input: 100-240V~ or 100-277V~, 50-60Hz, 2.0A;  
Output: 5-56 VDC, Max.8.0A, Max 70W

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

GT\*96700-\*\*\*\*\*

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1

As shown in the Test Report Ref. No. which forms part of this Certificate

171201719SHA-001

This CB Test Certificate is issued by the National Certification Body

**Intertek Semko AB**  
Box 1103  
SE-164 22 Kista, Sweden  
Int +46 8 750 00 00

Date: 18 June 2018

**intertek**

Signature:

Gary Hu



**Factories**

Globtek, Inc.  
186 Veterans Drive Northvale, NJ 07647 U.S.A.

GlobTek (Suzhou) Co., Ltd.  
Building 4, No.76, Jin Ling East Road, Suzhou Industrial Park, Suzhou, JiangSu, 215021, P. R. China

**Ratings and principal characteristics**

Model	Input	Output Voltage	Max. output current	Max. output power
GT*96700-*****	100-240V~ or 100-277V~, 50-60Hz,2.0A	5-8.9Vdc	8A	49W
	100-240V~ or 100-277V~, 50-60Hz,2.0A	9-44Vdc	5.42A	65W
	100-240V~ or 100-277V~, 50-60Hz,2.0A	44.1-56Vdc	1.48A	70W

**Additional information (if necessary)**

Explanation of model GT\*96700-\*\*\*\*\*:

The 1<sup>st</sup> symbol "\*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2<sup>nd</sup> symbol "\*" can be B, or C and is related to PCB size: B=2"x4", C=3"x5"

The 3<sup>rd</sup> symbol "\*" denotes the rated output wattage designation, which can be "01" to "70", with interval of 1.

The 4<sup>th</sup> symbol "\*" denote the standard rated output voltage designation, which can be "05" to "56" or "5.0" to "56.0" in 0.1V increments.

The 5<sup>th</sup> symbol "\*" can be -F or -FW, -F means Class I, -FW means Class II

The last symbol "\*" denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

Group differences for the CENELEC countries, and national differences for Canada, USA and Switzerland have been checked.

Date: 18 June 2018

Signature: 