

Handled by  
Candy Liang  
Direct telephone  
+86 21 61278415  
Reference  
1812779  
E-mail  
Shanghai-LC@intertek.com  
Your Reference  
171201721HA

**Intertek S Mark Certificate with No. 1812779**

We are pleased to enclose the Intertek S Mark Certificate you have applied for.

**Thank You for S Marking your Products!**

You have taken the first step to S Mark, but have you taken the next step by letting your customers know about your investment? When you use the S Mark in your marketing you are making clear that your company is truly committed to quality and safety. Positioning and differentiate yourself as a company which is dedicated to safety enables you to increase your sales – because in Europe safety sells.

The S Mark is highly recognized among retailers and authorities since 1926.

The S Mark also strengthens your brand by signaling that safety is important to your company and that you care about the safety of your clients. Since we value you as a customer we want to make sure you are getting the best value out of the S Mark. Many of our customers are already using the S Mark in their marketing. Popular media to use the mark is in campaigns, on packaging and in product catalogues. To find out more and to download tools, please visit <http://www.intertek.com/marks/s/>.



If you have any questions don't hesitate to contact us at [info-sweden@intertek.com](mailto:info-sweden@intertek.com)

**Additional part of the certification**

In addition to your product certification, we are conducting inspections at your manufacturing site/s. The inspections are carried out by our inspectors or subcontractors. For details of the technical requirements for these inspections, please visit <http://intertek.com/europeaninspection>

Your success is important to us!

Best Regards,



Gary Hu  
Local Certification Manager

# Certificate



for European Product Safety

Reference No. 1812779

## Medical Power supply

*Type designation* GT\*96700-\*\*\*\*\*


*Certificate holder* **GlobTek, Inc.**  
**186 Veterans Drive Northvale, NJ 07647,**  
**USA**

*The product complies with the standard(s)* EN 60601-1: 2006+ A1: 2013+A12: 2014

*Date of expiry* 06 July 2023

Additional information in Appendix.

*Certification Body* Intertek Semko AB, Product Certification *Place* Shanghai

*Signed*  \_\_\_\_\_ *Date* 06 July 2018  
Gary Hu *Page 1 of 2*

*This certificate is issued in accordance with the terms and conditions set out in the Appendix.*



# Intertek Appendix

Reference No.1812779

*Technical data*

Type designation

GT\*96700-\*\*\*\*\*

Input

100-240V~ or 100-277V~, 50-60Hz, 2.0A

Output

5-56 VDC, Max.8.0A, Max 70W

BRAND name:



*Product information*

GT\*96700-\*\*\*\*\*:

The 1st "\*" part can be 'M' or '-' or 'H' for market identification and not related to safety.

The 2nd "\*" can be B, or C and is related to PCB size:, B=2"x4", C=3"x5"

The 3rd "\*" denotes the rated output wattage designation, which can be "01" to "70", with interval of 1.

The 4th "\*" denote the standard rated output voltage designation, which can be "05" to "56" or "5.0" to "56.0" in 0. 1V increments.

The 5th "\*" can be -F or -FW, -F means Class I, -FW means Class II

The last "\*" denote any six character = 0-9 or A-Z or ()[] or - or blank for marketing purposes.

*Manufacturing site(s)*

GlobTek (Suzhou) Co., Ltd  
Building 4, No. 76, JinLing East Road Suzhou Industrial Park,  
Suzhou JiangSu, 215121,  
P.R.China

Safety related parts of the user instructions shall be written in a language acceptable to the country where the product is to be used.

If the product is to be marketed in different European countries, it must be provided with a certified plug appropriate for each country.

*Certification Body*

Intertek Semko AB, Product Certification

Place Shanghai

*Signed*

Gary Hu

Date 06 July 2018

*This Certificate is for the exclusive use of Intertek's Client and is provided pursuant to the Certification agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this certificate. Only the Client is authorized to permit copying or distribution of this certificate and then only in its entirety. Use of Intertek's Certification mark is restricted to the conditions laid out in the agreement. Any further use of the Intertek name for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. Initial Factory Assessments and Follow up Services are for the purpose of assuring appropriate usage of the Certification mark in accordance with the agreement, they are not for the purposes of production quality control and do not relieve the Client of their obligations in this respect.*